

EQA round: RF2/22 - Diagnostics of Rheumatoid Arthritis and ASLO

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

| Sample | Group | N | Result | N _{rel} |
|--------|-------|---|--------|------------------|
|--------|-------|---|--------|------------------|

Set 1

(231) Rheumat. fac. (scr)

| | | | | |
|---------------------------|-------------|----------------|----------|--------|
| A1 | All results | 1 | Negative | 0,59 % |
| | | >>> 169 | Positive | 99 % |
| Total per sample: n = 170 | | Success: 99 % | | |
| B1 | All results | >>> 170 | Negative | 100 % |
| Total per sample: n = 170 | | Success: 100 % | | |
| Total success: 99 % | | | | |

(233) Rheumat. factor IgG

| | | | | |
|--------------------------|-------------|----------------|----------------------------------|-------|
| A1 | All results | >>> 48 | Positive | 100 % |
| Total per sample: n = 48 | | Success: 100 % | | |
| B1 | All results | >>> 47 | Negative | 98 % |
| | | 1 | Inconclusive result (borderline) | 2,1 % |
| Total per sample: n = 48 | | Success: 98 % | | |
| Total success: 98 % | | | | |

(234) Rheumat. factor IgA

| | | | | |
|--------------------------|-------------|----------------|----------|-------|
| A1 | All results | >>> 46 | Positive | 100 % |
| Total per sample: n = 46 | | Success: 100 % | | |
| B1 | All results | >>> 46 | Negative | 100 % |
| Total per sample: n = 46 | | Success: 100 % | | |
| Total success: 100 % | | | | |

(235) Rheumat. factor IgM

| | | | | |
|--------------------------|-------------|----------------|----------|-------|
| A1 | All results | >>> 46 | Positive | 100 % |
| Total per sample: n = 46 | | Success: 100 % | | |
| B1 | All results | >>> 46 | Negative | 100 % |
| Total per sample: n = 46 | | Success: 100 % | | |
| Total success: 100 % | | | | |

(277) anti-CCP IgG

| | | | | |
|--------------------------|-------------|----------------|----------|-------|
| A1 | All results | >>> 82 | Positive | 100 % |
| Total per sample: n = 82 | | Success: 100 % | | |
| B1 | All results | >>> 82 | Negative | 100 % |
| Total per sample: n = 82 | | Success: 100 % | | |
| Total success: 100 % | | | | |

(282) anti-MCV

| | | | | |
|--------------------------|-------------|----------------|----------|-------|
| A1 | All results | >>> 23 | Positive | 100 % |
| Total per sample: n = 23 | | Success: 100 % | | |
| B1 | All results | >>> 23 | Negative | 100 % |
| Total per sample: n = 23 | | Success: 100 % | | |
| Total success: 100 % | | | | |