

**Summary statistics - quantitative results****EQA round: DD1/23 - D Dimers**

Deadline: 31.3.2023

Setup: groups - manufacturer of kit; Slovakia; minimal size of the groups n = 5

RoM = robust average	AV = assigned value	D <sub>max</sub> = acceptable difference
SD = standard deviation	CVPG = consensus of the participants' groups	LL = lower limit
CV = coefficient of variation	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	UL = upper limit
N <sub>tot</sub> = total number of the results		N <sub>eva</sub> = number of the results assessed
N <sub>out</sub> = number of the results removed before calculation		N <sub>suc</sub> = number of successful results
		S <sub>rel</sub> = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability							
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
(476) D Dimers (FEU)	[mg/L FEU]				56							56	55	98
<b>Sample A</b>												56	55	98
(917) Siemens Innovance D-Dimer	0,666	0,26	39	56			CVPG 0,861	0,007	27%	0,628	1,1		28	
(935) Mindray D-Dimer	0,864	0,086	9,9	28	0		CVPG 0,398	0,038 ± 0,189	0,209	0,587			15	
(976) IL HemosIL D-Dimer HS 500	0,398	0,06	15	15	0		CVPG 0,753	0,013	27%	0,549	0,957		6	
(988) Diagon Dia-D-Dimer	0,738	0,053	7,1	6	0		CVPG 0,455	0,068 ± 0,189	0,266	0,644			6	
Other	0,455	0,049	11	6	0		CVPG 0,455	0,068 ± 0,189	0,266	0,644			1	
					1	0								1
							1x 955							
<b>Sample B</b>												56	55	98
(917) Siemens Innovance D-Dimer	0,31	0,11	34	56			CVPG 0,358	0,004 ± 0,189	0,169	0,547			28	
(935) Mindray D-Dimer	0,367	0,048	13	28	0		CVPG 0,218	0,014 ± 0,189	0,029	0,407			15	
(976) IL HemosIL D-Dimer HS 500	0,218	0,022	10	15	0		CVPG 0,393	0,008 ± 0,189	0,204	0,582			6	
(988) Diagon Dia-D-Dimer	0,421	0,087	21	6	0		CVPG 0,203	0,022 ± 0,189	0,014	0,392			6	
Other	0,203	0,016	7,7	6	0		CVPG 0,203	0,022 ± 0,189	0,014	0,392			1	
					1	0								1
							1x 955							