

EQA round: FC2/23 - Calprotectin in Stool

Deadline: 22.9.2023

Setup: all data

AV = assigned value

CVP = consensus of all participants

>>> ... expected result

> ... acceptable result

± ... result not assessed

N_{tot} = total number of the resultsN_{rel} = relative number of the resultsN_{eva} = number of the results assessedN_{suc} = number of successful resultsS_{rel} = relative success

Test Sample Group	Frequency of the results				Success		
	AV	N _{tot}	N _{rel} [%]	Result	N _{eva}	N _{suc}	S _{rel} [%]
(223) Calprotectin		72			72	71	99
Sample A		72			72	71	99
All results	CVP >>>	71	99	Negative			
		1	1,4	Positive			
Sample B		72			72	72	100
All results	CVP >>>	72	100	Positive			