

## EQA round: FC2/23 - Calprotectin in Stool

Deadline: 22.9.2023

Setup: groups - manufacturer of kit; minimal size of the groups n = 5

RoM = robust average

SD = standard deviation

CV = coefficient of variation

N<sub>tot</sub> = total number of the resultsN<sub>out</sub> = number of the results removed before calculation

AV = assigned value

CVP = consensus of all participants

U<sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)D<sub>max</sub> = acceptable difference

LL = lower limit

UL = upper limit

N<sub>eva</sub> = number of the results assessedN<sub>suc</sub> = number of successful resultsS<sub>rel</sub> = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability						N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL					
<b>(223) Calprotectin</b>	[µg/g]				72								72	29	40	
<b>Sample A</b>		20	12	58	72		CVP	20	3,6	50%	10	30	72	45	63	
(135) OrgenTec		11,7	0,96	8,3	6	1							6			
(164) DiaSorin		5,5	0,74	13	6	0							6			
(166) Thermo Fisher		7	3	42	8	4							8			
(211) INOVA Diagnostics		17	1,5	8,7	6	0							6			
(219) Bühlmann		27,6	5,3	19	26	0							26			
(999) another manufacturer		19,3	11	59	14	3							14			
Other					6	1							6			
							2x 49, 1x 60, 1x 87, 1x 199, 1x 200									
<b>Sample B</b>		493	230	47	72		CVP	493	68	50%	246	740	72	50	69	
(135) OrgenTec		990	15	1,5	6	2							6			
(164) DiaSorin		334	17	5,1	6	0							6			
(166) Thermo Fisher		652	330	51	8	0							8			
(211) INOVA Diagnostics		253	58	23	6	0							6			
(219) Bühlmann		543	190	35	26	1							26			
(999) another manufacturer		410	190	47	14	1							14			
Other					6	0							6			
							2x 49, 1x 60, 1x 87, 1x 199, 1x 200									