

## EQA round: HS2/23 - Haemocoagulation Special

Deadline: 17.11.2023

Setup: groups - R (manufacturer of kit)

AV = assigned value

N<sub>tot</sub> = total number of the resultsN<sub>eva</sub> = number of the results assessed

CVP = consensus of all participants

N<sub>rel</sub> = relative number of the resultsN<sub>suc</sub> = number of successful results

&gt;&gt;&gt; ... expected result

S<sub>rel</sub> = relative success

&gt; ... acceptable result

± ... result not assessed

Test Sample Group	Frequency of the results					Success		
	AV	N <sub>tot</sub>	N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	
<b>(488) APC - resistance</b>		61			61	59	97	
<b>Sample A3</b>		61			61	59	97	
(38) IL		1	1,6	Negative				
	CVP >>>	14	23	Positive				
(51) Technoclone	CVP >>>	1	1,6	Positive				
(94) Stago	CVP >>>	3	4,9	Positive				
(102) Chromogenix	CVP >>>	1	1,6	Positive				
(149) Siemens (Dade, BN, Dimension)	CVP >>>	10	16	Positive				
(179) Siemens	CVP >>>	3	4,9	Positive				
(192) Pentapharm	CVP >>>	27	44	Positive				
(193) HYPHEN BioMed		1	1,6	Inconclusive result (borderline)				