

EQA round: IFT2/23 - Immunophenotypisation

Deadline: 30.10.2023

Setup: Slovakia

AV = assigned value

N_{tot} = total number of the resultsN_{eva} = number of the results assessed

CVP = consensus of all participants

N_{rel} = relative number of the resultsN_{suc} = number of successful results

>>> ... expected result

S_{rel} = relative success

> ... acceptable result

± ... result not assessed

Test Sample Group	Frequency of the results				Success		
	AV	N _{tot}	N _{rel} [%]	Result	N _{eva}	N _{suc}	S _{rel} [%]
(259) HLA-B27		20			20	16	80
Sample A		20			20	20	100
All results							
	>	1	5	Inconclusive result (borderline)			
	CVP >>>	19	95	Positive			
Sample B		20			20	16	80
All results							
	CVP >>>	16	80	Negative			
		3	15	Inconclusive result (borderline)			
		1	5	Positive			