

EQA round: KD3/23 - Glycated Haemoglobin

Deadline: 28.7.2023

Setup: groups - manufacturer of kit; minimal size of the groups n = 5

RoM = robust average

SD = standard deviation

CV = coefficient of variation

N_{tot} = total number of the resultsN_{out} = number of the results removed before calculation

AV = assigned value

CVP = consensus of all participants

U_{AV} = expanded uncertainty of the assigned value (k = 2)D_{max} = acceptable difference

LL = lower limit

UL = upper limit

N_{eva} = number of the results assessedN_{suc} = number of successful resultsS_{rel} = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability						N _{eva}	N _{suc}	S _{rel} [%]
							AV	U _{AV}	D _{max}	LL	UL				
(220) Haemoglobin A _{1c}	[mmol/mol]				141								137	134	98
Sample A		32,8	1,6	4,8	141								137	135	99
(14) Bio-Rad		33	1,7	5	41	0	CVP	32,8	0,32	12%	28,8	36,8	41		
(25) TOSOH		33,5	1	3	44	0	CVP	32,8	0,32	12%	28,8	36,8	44		
(60) Roche		32,1	2,9	9,2	8	0	CVP	32,8	0,32	12%	28,8	36,8	8		
(150) Arkray		31,8	1,2	3,9	37	0	CVP	32,8	0,32	12%	28,8	36,8	37		
Other					11	0							7		
								3x 1, 1x 12, 3x 34, 1x 46, 1x 58, 1x 149, 1x 162							
Sample B		74,5	2,9	3,9	141								137	136	99
(14) Bio-Rad		76,1	1,9	2,5	41	0	CVP	74,6	0,58	12%	65,6	83,6	41		
(25) TOSOH		74,9	1,5	2	44	0	CVP	74,6	0,58	12%	65,6	83,6	44		
(60) Roche		76,9	2	2,6	8	0	CVP	74,6	0,58	12%	65,6	83,6	8		
(150) Arkray		71,4	2,5	3,5	37	0	CVP	74,6	0,58	12%	65,6	83,6	37		
Other					11	0							7		
								3x 1, 1x 12, 3x 34, 1x 46, 1x 58, 1x 149, 1x 162							