

**Summary statistics - qualitative results****EQA round: RF1/23 - Diagnostics of Rheumatoid Arthritis and ASLO**

Deadline: 17.3.2023

Setup: groups - manufacturer of kit

AV = assigned value  
 CVP = consensus of all participants  
 >>> ... expected result  
 > ... acceptable result  
 ± ... result not assessed

N<sub>tot</sub> = total number of the results  
 N<sub>rel</sub> = relative number of the results

N<sub>eva</sub> = number of the results assessed  
 N<sub>suc</sub> = number of successful results  
 S<sub>rel</sub> = relative success

Test Sample Group	AV	N <sub>tot</sub>	N <sub>rel</sub> [%]	Frequency of the results		Success		
				Result		N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
(277) anti-CCP IgG		83				83	83	100
<b>Sample A1</b>		83				83	83	100
(1) Abbott	CVP	>>>	20	24	Positive			
(14) Bio-Rad	CVP	>>>	1	1,2	Positive			
(29) Siemens (Immulite)	CVP	>>>	5	6	Positive			
(60) Roche	CVP	>>>	12	14	Positive			
(66) DRG	CVP	>>>	1	1,2	Positive			
(73) Snibe	CVP	>>>	1	1,2	Positive			
(87) Euroimmun	CVP	>>>	9	11	Positive			
(135) OrgenTec	CVP	>>>	3	3,6	Positive			
(140) TestLine	CVP	>>>	9	11	Positive			
(147) AIDA	CVP	>>>	1	1,2	Positive			
(148) Aesku.lab Diagnostika	CVP	>>>	2	2,4	Positive			
(162) Siemens (Atellica)	CVP	>>>	3	3,6	Positive			
(166) Thermo Fisher	CVP	>>>	7	8,4	Positive			
(179) Siemens	CVP	>>>	1	1,2	Positive			
(211) INOVA Diagnostics	CVP	>>>	5	6	Positive			
(999) another manufacturer	CVP	>>>	3	3,6	Positive			
<b>Sample B1</b>		83				83	83	100
(1) Abbott	CVP	>>>	20	24	Negative			
(14) Bio-Rad	CVP	>>>	1	1,2	Negative			
(29) Siemens (Immulite)	CVP	>>>	5	6	Negative			
(60) Roche	CVP	>>>	12	14	Negative			
(66) DRG	CVP	>>>	1	1,2	Negative			
(73) Snibe	CVP	>>>	1	1,2	Negative			
(87) Euroimmun	CVP	>>>	9	11	Negative			
(135) OrgenTec	CVP	>>>	3	3,6	Negative			
(140) TestLine	CVP	>>>	9	11	Negative			
(147) AIDA	CVP	>>>	1	1,2	Negative			
(148) Aesku.lab Diagnostika	CVP	>>>	2	2,4	Negative			
(162) Siemens (Atellica)	CVP	>>>	3	3,6	Negative			
(166) Thermo Fisher	CVP	>>>	7	8,4	Negative			
(179) Siemens	CVP	>>>	1	1,2	Negative			
(211) INOVA Diagnostics	CVP	>>>	5	6	Negative			
(999) another manufacturer	CVP	>>>	3	3,6	Negative			
(282) anti-MCV		23				23	23	100
<b>Sample A1</b>		23				23	23	100
(135) OrgenTec	CVP	>>>	22	96	Positive			
(166) Thermo Fisher	CVP	>>>	1	4,3	Positive			
<b>Sample B1</b>		23				23	23	100
(135) OrgenTec	CVP	>>>	22	96	Negative			
(166) Thermo Fisher	CVP	>>>	1	4,3	Negative			
(231) Rheumat. fac. (scr)		173				173	173	100
<b>Sample A1</b>		173				173	173	100
(1) Abbott	CVP	>>>	39	23	Positive			
(12) Beckman Coulter	CVP	>>>	5	2,9	Positive			
(58) Beckman Coulter (AU)	CVP	>>>	26	15	Positive			
(60) Roche	CVP	>>>	50	29	Positive			
(70) Wako	CVP	>>>	1	0,6	Positive			
(91) Binding Site	CVP	>>>	3	1,7	Positive			
(115) BioSystems	CVP	>>>	3	1,7	Positive			
(116) Sentinel Diagnostics	CVP	>>>	2	1,2	Positive			
(148) Aesku.lab Diagnostika	CVP	>>>	2	1,2	Positive			
(149) Siemens (Dade, BN, Dimension)	CVP	>>>	11	6,4	Positive			
(162) Siemens (Atellica)	CVP	>>>	13	7,5	Positive			
(177) Mindray	CVP	>>>	5	2,9	Positive			
(178) DiaSys	CVP	>>>	3	1,7	Positive			
(179) Siemens	CVP	>>>	9	5,2	Positive			
(999) another manufacturer	CVP	>>>	1	0,6	Positive			
<b>Sample B1</b>		173				173	173	100
(1) Abbott	CVP	>>>	39	23	Negative			
(12) Beckman Coulter	CVP	>>>	5	2,9	Negative			
(58) Beckman Coulter (AU)	CVP	>>>	26	15	Negative			
(60) Roche	CVP	>>>	50	29	Negative			

**Summary statistics - qualitative results****EQA round: RF1/23 - Diagnostics of Rheumatoid Arthritis and ASLO**

Deadline: 17.3.2023

Setup: groups - manufacturer of kit

Test Sample Group	AV	N <sub>tot</sub>	N <sub>rel</sub> [%]	Frequency of the results		Success N <sub>eva</sub> N <sub>suc</sub> S <sub>rel</sub> [%]
				Result		
(70) Wako	CVP >>>	1	0,6	Negative		
(91) Binding Site	CVP >>>	3	1,7	Negative		
(115) BioSystems	CVP >>>	3	1,7	Negative		
(116) Sentinel Diagnostics	CVP >>>	2	1,2	Negative		
(148) Aesku.lab Diagnostika	CVP >>>	2	1,2	Negative		
(149) Siemens (Dade, BN, Dimension)	CVP >>>	11	6,4	Negative		
(162) Siemens (Atellica)	CVP >>>	13	7,5	Negative		
(177) Mindray	CVP >>>	5	2,9	Negative		
(178) DiaSys	CVP >>>	3	1,7	Negative		
(179) Siemens	CVP >>>	9	5,2	Negative		
(999) another manufacturer	CVP >>>	1	0,6	Negative		
<b>(234) Rheumat. factor IgA</b>		<b>47</b>				<b>47 47 100</b>
<b>Sample A1</b>		<b>47</b>				<b>47 47 100</b>
(5) Dialab	CVP >>>	1	2,1	Positive		
(135) OrgenTec	CVP >>>	12	26	Positive		
(140) TestLine	CVP >>>	21	45	Positive		
(147) AIDA	CVP >>>	2	4,3	Positive		
(148) Aesku.lab Diagnostika	CVP >>>	4	8,5	Positive		
(166) Thermo Fisher	CVP >>>	6	13	Positive		
(211) INOVA Diagnostics	CVP >>>	1	2,1	Positive		
<b>Sample B1</b>		<b>47</b>				<b>47 47 100</b>
(5) Dialab	CVP >>>	1	2,1	Negative		
(135) OrgenTec	CVP >>>	12	26	Negative		
(140) TestLine	CVP >>>	21	45	Negative		
(147) AIDA	CVP >>>	2	4,3	Negative		
(148) Aesku.lab Diagnostika	CVP >>>	4	8,5	Negative		
(166) Thermo Fisher	CVP >>>	6	13	Negative		
(211) INOVA Diagnostics	CVP >>>	1	2,1	Negative		
<b>(233) Rheumat. factor IgG</b>		<b>49</b>				<b>49 45 92</b>
<b>Sample A1</b>		<b>49</b>				<b>49 45 92</b>
(1) Abbott	CVP >>>	1	2	Positive		
(5) Dialab	CVP >>>	1	2	Positive		
(58) Beckman Coulter (AU)	CVP >>>	1	2	Positive		
(135) OrgenTec	CVP >>>	12	24	Positive		
(140) TestLine	CVP >>>	21	43	Positive		
(147) AIDA	CVP >>>	2	4,1	Positive		
(148) Aesku.lab Diagnostika	CVP >>>	4	8,2	Positive		
(166) Thermo Fisher	CVP >>>	4	8,2	Negative		
	CVP >>>	2	4,1	Positive		
(211) INOVA Diagnostics	CVP >>>	1	2	Positive		
<b>Sample B1</b>		<b>49</b>				<b>49 49 100</b>
(1) Abbott	CVP >>>	1	2	Negative		
(5) Dialab	CVP >>>	1	2	Negative		
(58) Beckman Coulter (AU)	CVP >>>	1	2	Negative		
(135) OrgenTec	CVP >>>	12	24	Negative		
(140) TestLine	CVP >>>	21	43	Negative		
(147) AIDA	CVP >>>	2	4,1	Negative		
(148) Aesku.lab Diagnostika	CVP >>>	4	8,2	Negative		
(166) Thermo Fisher	CVP >>>	6	12	Negative		
(211) INOVA Diagnostics	CVP >>>	1	2	Negative		
<b>(235) Rheumat. factor IgM</b>		<b>47</b>				<b>47 47 100</b>
<b>Sample A1</b>		<b>47</b>				<b>47 47 100</b>
(5) Dialab	CVP >>>	1	2,1	Positive		
(135) OrgenTec	CVP >>>	12	26	Positive		
(140) TestLine	CVP >>>	21	45	Positive		
(147) AIDA	CVP >>>	2	4,3	Positive		
(148) Aesku.lab Diagnostika	CVP >>>	4	8,5	Positive		
(166) Thermo Fisher	CVP >>>	6	13	Positive		
(211) INOVA Diagnostics	CVP >>>	1	2,1	Positive		
<b>Sample B1</b>		<b>47</b>				<b>47 47 100</b>
(5) Dialab	CVP >>>	1	2,1	Negative		
(135) OrgenTec	CVP >>>	12	26	Negative		
(140) TestLine	CVP >>>	21	45	Negative		
(147) AIDA	CVP >>>	2	4,3	Negative		
(148) Aesku.lab Diagnostika	CVP >>>	4	8,5	Negative		
(166) Thermo Fisher	CVP >>>	6	13	Negative		
(211) INOVA Diagnostics	CVP >>>	1	2,1	Negative		