

Summary statistics - quantitative results**EQA round: RF1/23 - Diagnostics of Rheumatoid Arthritis and ASLO**

Deadline: 17.3.2023

Setup: groups - manufacturer of kit; minimal size of the groups n = 5

RoM = robust average
 SD = standard deviation
 CV = coefficient of variation
 N_{tot} = total number of the results
 N_{out} = number of the results removed before calculation

AV = assigned value
 CVPG = consensus of the participants' groups
 U_{AV} = expanded uncertainty of the assigned value ($k = 2$)

D_{max} = acceptable difference
 LL = lower limit
 UL = upper limit
 N_{eva} = number of the results assessed
 N_{suc} = number of successful results
 S_{rel} = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N_{tot}	N_{out}	Comparability						N_{eva}	N_{suc}	S_{rel} [%]
							AV	U_{AV}	D_{max}	LL	UL				
(237) anti-streptolysin O (ASLO)	[kU/L]				167								148	141	95
Sample A2													148	145	98
(1) Abbott		309	63	20	167										
(58) Beckman Coulter (AU)		228	6,7	2,9	42	0	CVPG	228	2,5	20%	182	274			42
(60) Roche		319	18	5,6	30	0	CVPG	319	8	20%	255	383			30
(149) Siemens (Dade, BN, Dimension)		345	16	4,7	46	0	CVPG	345	5,8	20%	276	414			46
(162) Siemens (Atellica)		330	16	4,7	9	0	CVPG	330	19	20%	264	396			9
(179) Siemens		371	15	4,1	13	0	CVPG	371	10	20%	296	446			13
Other		371	24	6,4	8	0	CVPG	371	18	20%	296	446			8
					19	0									0
							4x 12, 3x 91, 3x 115, 1x 116, 4x 177, 3x 178, 1x 999								
Sample B2		77	13	17	167								148	142	96
(1) Abbott		61,5	3,7	6	42	0	CVPG	61,5	1,4	20%	49,2	73,8			42
(58) Beckman Coulter (AU)		79,3	8,1	10	30	0	CVPG	79,3	3,6	20%	63,4	95,2			30
(60) Roche		82,4	7,1	8,6	46	0	CVPG	82,4	2,6	20%	65,9	98,9			46
(149) Siemens (Dade, BN, Dimension)		88	4,4	5,1	9	0	CVPG	88	5,5	20%	70,4	106			9
(162) Siemens (Atellica)		83,5	5,4	6,4	13	0	CVPG	83,5	3,7	20%	66,8	101			13
(179) Siemens		86,5	2,2	2,6	8	0	CVPG	86,5	1,7	20%	69,2	104			8
Other					19	0									0
							4x 12, 3x 91, 3x 115, 1x 116, 4x 177, 3x 178, 1x 999								