

Summary statistics - qualitative results**EQA round: T23/23 - Antibodies IgG anti EM and anti tTG (study)**

Deadline: 23.10.2023

Setup: groups - measurement principle

AV = assigned value
 CVP = consensus of all participants
 >>> ... expected result
 > ... acceptable result
 ± ... result not assessed

N_{tot} = total number of the results
 N_{rel} = relative number of the results

N_{eva} = number of the results assessed
 N_{suc} = number of successful results
 S_{rel} = relative success

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success		
				Result		N _{eva}	N _{suc}	S _{rel} [%]
(437) anti-endomysium IgG		51				51	45	88
Sample A		51				51	49	96
(0) Not specified	CVP >>>	6	12	Negative				
		1	2	Inconclusive result (borderline)				
(1) Fluorescent methods	CVP >>>	43	84	Negative				
		1	2	Positive				
Sample B		51				51	47	92
(0) Not specified	CVP >>>	7	14	Positive				
(1) Fluorescent methods		2	3,9	Negative				
		2	3,9	Inconclusive result (borderline)				
	CVP >>>	40	78	Positive				
(438) anti-transglutaminase IgG		74				74	73	99
Sample A		74				74	73	99
(0) Not specified	CVP >>>	7	9,5	Negative				
(1) Fluorescent methods	CVP >>>	1	1,4	Negative				
(2) EIA methods	CVP >>>	51	69	Negative				
		1	1,4	Positive				
(5) LIA, ILMA	CVP >>>	11	15	Negative				
(99) Another measurement principle	CVP >>>	3	4,1	Negative				
Sample B		74				74	73	99
(0) Not specified	CVP >>>	7	9,5	Positive				
(1) Fluorescent methods	CVP >>>	1	1,4	Positive				
(2) EIA methods		1	1,4	Negative				
	CVP >>>	51	69	Positive				
(5) LIA, ILMA	CVP >>>	11	15	Positive				
(99) Another measurement principle	CVP >>>	3	4,1	Positive				