

Summary statistics - qualitative results**EQA round: T23/23 - Antibodies IgG anti EM and anti tTG (study)**

Deadline: 23.10.2023

Setup: groups - manufacturer of kit

AV = assigned value
 CVP = consensus of all participants
 >>> ... expected result
 > ... acceptable result
 ± ... result not assessed

N_{tot} = total number of the results
 N_{rel} = relative number of the results

N_{eva} = number of the results assessed
 N_{suc} = number of successful results
 S_{rel} = relative success

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(437) anti-endomysium IgG		51				51	45	88
Sample A		51				51	49	96
(0) not specified	CVP >>>	6	12	Negative				
(87) Euroimmun	CVP >>>	16	31	Negative				
		1	2	Inconclusive result (borderline)				
		1	2	Positive				
(115) BioSystems	CVP >>>	6	12	Negative				
(140) TestLine	CVP >>>	1	2	Negative				
(148) Aesku.lab Diagnostika	CVP >>>	4	7,8	Negative				
(211) INOVA Diagnostics	CVP >>>	15	29	Negative				
(998) home-made	CVP >>>	1	2	Negative				
Sample B		51				51	47	92
(0) not specified	CVP >>>	6	12	Positive				
(87) Euroimmun		2	3,9	Negative				
		1	2	Inconclusive result (borderline)				
(115) BioSystems	CVP >>>	15	29	Positive				
(140) TestLine	CVP >>>	6	12	Positive				
(148) Aesku.lab Diagnostika	CVP >>>	1	2	Positive				
(211) INOVA Diagnostics	CVP >>>	4	7,8	Positive				
(998) home-made	CVP >>>	15	29	Positive				
		1	2	Inconclusive result (borderline)				
(438) anti-transglutaminase IgG		74				74	73	99
Sample A		74				74	73	99
(0) not specified	CVP >>>	7	9,5	Negative				
(8) Eurospital	CVP >>>	2	2,7	Negative				
(51) Technoclone	CVP >>>	1	1,4	Negative				
(52) Dynex	CVP >>>	1	1,4	Negative				
(87) Euroimmun	CVP >>>	6	8,1	Negative				
(115) BioSystems	CVP >>>	3	4,1	Negative				
(135) OrgenTec	CVP >>>	7	9,5	Negative				
		1	1,4	Positive				
(140) TestLine	CVP >>>	16	22	Negative				
(148) Aesku.lab Diagnostika	CVP >>>	4	5,4	Negative				
(166) Thermo Fisher	CVP >>>	6	8,1	Negative				
(183) Immunodiagnostic Systems	CVP >>>	2	2,7	Negative				
(191) D-tek	CVP >>>	1	1,4	Negative				
(199) Eurolab Lambda	CVP >>>	1	1,4	Negative				
(211) INOVA Diagnostics	CVP >>>	14	19	Negative				
(999) another manufacturer	CVP >>>	2	2,7	Negative				
Sample B		74				74	73	99
(0) not specified	CVP >>>	7	9,5	Positive				
(8) Eurospital	CVP >>>	2	2,7	Positive				
(51) Technoclone	CVP >>>	1	1,4	Positive				
(52) Dynex	CVP >>>	1	1,4	Positive				
(87) Euroimmun	CVP >>>	6	8,1	Positive				
(115) BioSystems	CVP >>>	3	4,1	Positive				
(135) OrgenTec		1	1,4	Negative				
	CVP >>>	7	9,5	Positive				
(140) TestLine	CVP >>>	16	22	Positive				
(148) Aesku.lab Diagnostika	CVP >>>	4	5,4	Positive				
(166) Thermo Fisher	CVP >>>	6	8,1	Positive				
(183) Immunodiagnostic Systems	CVP >>>	2	2,7	Positive				
(191) D-tek	CVP >>>	1	1,4	Positive				
(199) Eurolab Lambda	CVP >>>	1	1,4	Positive				
(211) INOVA Diagnostics	CVP >>>	14	19	Positive				
(999) another manufacturer	CVP >>>	2	2,7	Positive				