

EQA round: AKS3/24 - Basic Clinical Chemistry - Serum

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

RoM = robust average	AV = assigned value	D _{max} = acceptable difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	CVP = consensus of all participants	UL = upper limit
N _{tot} = total number of the results	CVPG = consensus of the participants' groups	N _{eva} = number of the results assessed
N _{out} = number of the results removed before calculation	U _{AV} = expanded uncertainty of the assigned value (k = 2)	N _{suc} = number of successful results
		S _{rel} = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					Traceability								
							AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	AV	U _{AV}	D _{max}	LL	UL	N _{eva}
(22) α-amylase	[μkat/L]				150							0						148	144	97
Sample A		5,93	0,42	7	150							0						148	146	99
(1) Abbott		6,28	0,13	2,1	30	0							CRV	5,98	0,16	15%	5,08	6,88		30
(46) Erba Lachema		6,07	0,11	1,8	9	0							CRV	5,98	0,16	15%	5,08	6,88		9
(58) Beckman Coulter (AU)		6,01	0,22	3,6	22	0							CRV	5,98	0,16	15%	5,08	6,88		22
(60) Roche		5,57	0,1	1,9	62	0							CRV	5,98	0,16	15%	5,08	6,88		62
(162) Siemens (Atellica)		6,39	0,096	1,5	15	0							CRV	5,98	0,16	15%	5,08	6,88		15
Other					12	0						0								10
							2x 149							1x 12, 1x 49, 1x 85, 2x 149, 4x 177, 3x 179						
Sample B		4,85	0,36	7,4	150							0						148	144	97
(1) Abbott		5,19	0,11	2,2	30	0							CRV	4,72	0,14	15%	4,01	5,43		30
(46) Erba Lachema		4,95	0,16	3,1	9	0							CRV	4,72	0,14	15%	4,01	5,43		9
(58) Beckman Coulter (AU)		4,88	0,15	3	22	0							CRV	4,72	0,14	15%	4,01	5,43		22
(60) Roche		4,53	0,081	1,8	62	0							CRV	4,72	0,14	15%	4,01	5,43		62
(162) Siemens (Atellica)		5,23	0,083	1,6	15	0							CRV	4,72	0,14	15%	4,01	5,43		15
Other					12	0						0								10
							2x 149							1x 12, 1x 49, 1x 85, 2x 149, 4x 177, 3x 179						
(35) α-amylase pancreatic	[μkat/L]				67							67	66	99						0
Sample A		5,01	0,25	5	67		CVP	5,01	0,074	18%	4,1	5,92	67	66	99					0
(1) Abbott		4,87	0,2	4,1	15	0							15							
(46) Erba Lachema		5,25	0,27	5,1	8	0							8							
(60) Roche		4,91	0,14	2,8	27	0							27							
(162) Siemens (Atellica)		5,66	0,052	0,92	7	0							7							
Other					10	0							10							
							3x 58, 4x 116, 1x 178, 2x 179													
Sample B		4,13	0,18	4,3	67		CVP	4,13	0,053	18%	3,38	4,88	67	66	99					0
(1) Abbott		4,03	0,16	4,1	15	0							15							
(46) Erba Lachema		4,32	0,22	5	8	0							8							
(60) Roche		4,06	0,11	2,6	27	0							27							
(162) Siemens (Atellica)		4,69	0,045	0,95	7	0							7							
Other					10	0							10							
							3x 58, 4x 116, 1x 178, 2x 179													
(32) γ-globulin (elfo)	[-]				35							35	35	100						0
Sample A		0,127	0,014	11	35		CVP	0,127	0,006	30%	0,088	0,166	35	35	100					0
(34) Sebia		0,124	0,013	10	29	0							29							
(213) Interlab Srl		0,139	0,013	9,1	5	0							5							
Other					1	0							1							
							1x 0													
Sample B		0,123	0,012	9,8	35		CVP	0,123	0,005	30%	0,086	0,16	35	35	100					0
(34) Sebia		0,123	0,012	10	29	0							29							

EQA round: AKS3/24 - Basic Clinical Chemistry - Serum

Deadline: 26.7.2024

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]	Traceability					N _{eva}	N _{suc}	S _{rel} [%]
							AV	U _{AV}	D _{max}	LL	UL				AV	U _{AV}	D _{max}	LL	UL			
(60) Roche		0,726	0,052	7,1	60	0	CVPG	0,726	0,016	14%	0,624	0,828	60				60					
(162) Siemens (Atellica)		0,728	0,038	5,3	18	0	CVPG	0,728	0,022	14%	0,626	0,83	18				18					
Other					11	0							0				11					
							1x 12, 1x 85, 2x 149, 4x 177, 3x 179								1x 12, 1x 85, 2x 149, 4x 177, 3x 179							
(24) ALT	[μkat/L]				152							0					152	145	95			
Sample A		1,7	0,092	5,4	152							0	CRV	1,702	0,045	15%	1,44	1,96	152	150	99	
(1) Abbott		1,73	0,058	3,4	34	0													34			
(46) Erba Lachema		1,8	0,007	0,41	6	0													6			
(58) Beckman Coulter (AU)		1,7	0,051	3	21	0													21			
(60) Roche		1,63	0,049	3	62	0													62			
(162) Siemens (Atellica)		1,77	0,037	2,1	18	0													18			
Other					11	0													11			
							1x 12, 1x 85, 2x 149, 4x 177, 3x 179								1x 12, 1x 85, 2x 149, 4x 177, 3x 179							
Sample B		1,2	0,068	5,7	152							0	CRV	1,164	0,031	15%	0,989	1,34	152	145	95	
(1) Abbott		1,19	0,053	4,5	34	0													34			
(46) Erba Lachema		1,27	0,03	2,3	6	0													6			
(58) Beckman Coulter (AU)		1,21	0,048	4	21	0													21			
(60) Roche		1,16	0,033	2,9	62	0													62			
(162) Siemens (Atellica)		1,28	0,036	2,8	18	0													18			
Other					11	0													11			
							1x 12, 1x 85, 2x 149, 4x 177, 3x 179								1x 12, 1x 85, 2x 149, 4x 177, 3x 179							
(23) AST	[μkat/L]				152							0							152	149	98	
Sample A		1,48	0,069	4,6	152							0	CRV	1,473	0,044	15%	1,25	1,7	152	149	98	
(1) Abbott		1,52	0,039	2,6	34	0													34			
(46) Erba Lachema		1,45	0,045	3,1	6	0													6			
(58) Beckman Coulter (AU)		1,55	0,061	3,9	21	0													21			
(60) Roche		1,44	0,055	3,8	62	0													62			
(162) Siemens (Atellica)		1,49	0,038	2,5	18	0													18			
Other					11	0													11			
							1x 12, 1x 85, 2x 149, 4x 177, 3x 179								1x 12, 1x 85, 2x 149, 4x 177, 3x 179							
Sample B		2,03	0,082	4	152							0	CRV	2,017	0,044	15%	1,71	2,32	152	151	99	
(1) Abbott		2,04	0,047	2,3	34	0													34			
(46) Erba Lachema		1,99	0,037	1,9	6	0													6			
(58) Beckman Coulter (AU)		2,13	0,087	4,1	21	0													21			
(60) Roche		1,98	0,065	3,3	62	0													62			
(162) Siemens (Atellica)		2,07	0,033	1,6	18	0													18			
Other					11	0													11			
							1x 12, 1x 85, 2x 149, 4x 177, 3x 179								1x 12, 1x 85, 2x 149, 4x 177, 3x 179							
(13) Bilirubin total	[μmol/L]				152							0							152	152	100	
Sample A		21,7	2,3	11	152							0	CRV	21,42	0,57	26%	15,8	27	152	152	100	
(1) Abbott		22,6	1,4	6,1	34	0													34			
(46) Erba Lachema		23,3	0,22	0,96	5	0													5			
(58) Beckman Coulter (AU)		24,2	0,65	2,7	19	0													19			
(60) Roche		19,6	0,74	3,7	62	0													62			
(162) Siemens (Atellica)		23,4	1,4	5,9	18	0													18			

EQA round: AKS3/24 - Basic Clinical Chemistry - Serum

Deadline: 26.7.2024

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]	Traceability					N _{eva}	N _{suc}	S _{rel} [%]					
							AV	U _{AV}	D _{max}	LL	UL				AV	U _{AV}	D _{max}	LL	UL								
(12) Lactate	[mmol/L]				64							64	61	95											0		
Sample A		3,98	0,079	2	64		CVP	3,98	0,024	15%	3,38	4,58	64	64	100										0		
(1) Abbott		3,96	0,059	1,5	6	0						6															
(58) Beckman Coulter (AU)		3,96	0,037	0,94	9	0						9															
(60) Roche		3,98	0,064	1,6	25	0						25															
(162) Siemens (Atellica)		4,03	0,068	1,7	11	0						11															
Other					13	0						13															
								2x 12, 1x 49, 3x 61, 1x 75, 1x 115, 1x 116, 2x 149, 2x 179																			
Sample B		0,981	0,048	4,9	64		CVP	0,981	0,015	15%	0,833	1,13	64	61	95										0		
(1) Abbott		1,01	0,052	5,2	6	0						6															
(58) Beckman Coulter (AU)		0,935	0,022	2,4	9	0						9															
(60) Roche		0,985	0,028	2,9	25	0						25															
(162) Siemens (Atellica)		1,01	0,021	2	11	0						11															
Other					13	0						13															
								2x 12, 1x 49, 3x 61, 1x 75, 1x 115, 1x 116, 2x 149, 2x 179																			
(28) LD	[μkat/L]				120																				120	119	99
Sample A		3,36	0,13	3,8	120							0			CRV	3,368	0,086	18%	2,76	3,98				120	119	99	
(1) Abbott		3,29	0,2	6,1	24	0																			24		
(46) Erba Lachema		3,64	0,12	3,3	5	0																			5		
(58) Beckman Coulter (AU)		3,36	0,1	3,1	19	0																			19		
(60) Roche		3,34	0,09	2,7	48	0																			48		
(162) Siemens (Atellica)		3,37	0,085	2,5	14	0																			14		
Other					10	0																			10		
															1x 12, 1x 85, 2x 115, 1x 149, 2x 177, 3x 179												
Sample B		1,9	0,082	4,3	120							0			CRV	1,976	0,046	18%	1,62	2,34				120	120	100	
(1) Abbott		1,86	0,1	5,6	24	0																			24		
(46) Erba Lachema		2,01	0,11	5,5	5	0																			5		
(58) Beckman Coulter (AU)		1,88	0,081	4,3	19	0																			19		
(60) Roche		1,9	0,07	3,7	48	0																			48		
(162) Siemens (Atellica)		1,92	0,05	2,6	14	0																			14		
Other					10	0																			10		
															1x 12, 1x 85, 2x 115, 1x 149, 2x 177, 3x 179												
(29) Lipase	[μkat/L]				91							79	77	97												0	
Sample A		1,01	0,069	6,8	91							79	79	100												0	
(1) Abbott		1,04	0,05	4,7	16	0	CVPG	1,04	0,03	24%	0,79	1,29	16														
(58) Beckman Coulter (AU)		0,994	0,042	4,3	16	0	CVPG	0,994	0,026	24%	0,755	1,24	16														
(60) Roche		0,979	0,044	4,5	35	0	CVPG	0,979	0,018	24%	0,744	1,22	35														
(162) Siemens (Atellica)		1,09	0,032	3	12	0	CVPG	1,09	0,023	24%	0,828	1,36	12														
Other					12	0						0															
								1x 12, 1x 46, 2x 49, 1x 116, 2x 149, 1x 177, 2x 178, 2x 179																			
Sample B		0,507	0,034	6,7	91							79	77	97												0	
(1) Abbott		0,519	0,023	4,4	16	0	CVPG	0,519	0,014	24%	0,394	0,644	16														
(58) Beckman Coulter (AU)		0,465	0,038	8,1	16	0	CVPG	0,465	0,023	24%	0,353	0,577	16														
(60) Roche		0,505	0,022	4,3	35	0	CVPG	0,505	0,009	24%	0,383	0,627	35														
(162) Siemens (Atellica)		0,544	0,021	3,8	12	0	CVPG	0,544	0,015	24%	0,413	0,675	12														

