

Summary statistics - qualitative results**EQA round: DGP2/24 - Urine Strip Tests Analysis**

Deadline: 3.5.2024

Setup: groups - R (strip manufacturer)

AV = assigned value

N_{tot} = total number of the resultsN_{eva} = number of the results assessed

CVP = consensus of all participants

N_{rel} = relative number of the resultsN_{suc} = number of successful results

CVPG = consensus of the participants' groups

S_{rel} = relative success

>>> ... expected result

> ... acceptable result

± ... result not assessed

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(268) Bilirubin		69				64	63	98
Sample A		69				69	68	99
(5) Dialab	CVP	>>>	1	1,4 Negative				
(12) Beckman Coulter	CVP	>>>	4	5,8 Negative				
(18) Cormay	CVP	>>>	4	5,8 Negative				
			1	1,4 1+, 2+				
(46) Erba Lachema	CVP	>>>	10	14 Negative				
(49) BioVendor	CVP	>>>	1	1,4 Negative				
(60) Roche	CVP	>>>	4	5,8 Negative				
(63) Sysmex	CVP	>>>	4	5,8 Negative				
(109) Axis-Shield	CVP	>>>	1	1,4 Negative				
(162) Siemens (Atellica)	CVP	>>>	5	7,2 Negative				
(179) Siemens	CVP	>>>	5	7,2 Negative				
(201) Iris Diagnostics	CVP	>>>	1	1,4 Negative				
(202) Dirui	CVP	>>>	19	28 Negative				
(207) 77 Elektronika Kft.	CVP	>>>	6	8,7 Negative				
(999) another manufacturer	CVP	>>>	3	4,3 Negative				
Sample B		69				64	63	98
(5) Dialab	CVP	>>>	1	1,4 3+, 4+				
(12) Beckman Coulter; (223) Iris iChem Velocity		±	2	2,9 Negative				
		±	2	2,9 1+, 2+				
(18) Cormay			1	1,4 1+, 2+				
	CVP	>>>	4	5,8 3+, 4+				
(46) Erba Lachema	CVP	>>>	10	14 3+, 4+				
(49) BioVendor	CVP	>>>	1	1,4 3+, 4+				
(60) Roche	CVP	>>>	4	5,8 3+, 4+				
(63) Sysmex	CVP	>>>	4	5,8 3+, 4+				
(109) Axis-Shield	CVP	>>>	1	1,4 3+, 4+				
(162) Siemens (Atellica)	CVP	>>>	5	7,2 3+, 4+				
(179) Siemens	CVP	>>>	5	7,2 3+, 4+				
(201) Iris Diagnostics; (223) Iris iChem Velocity		±	1	1,4 Negative				
(202) Dirui	CVP	>>>	19	28 3+, 4+				
(207) 77 Elektronika Kft.	CVP	>>>	6	8,7 3+, 4+				
(999) another manufacturer	CVP	>>>	3	4,3 3+, 4+				
(264) Total protein		68				68	66	97
Sample A		68				68	68	100
(5) Dialab	CVP	>>>	1	1,5 Negative, < 0,3 g/L				
(12) Beckman Coulter	CVP	>>>	4	5,9 Negative, < 0,3 g/L				
(18) Cormay	CVP	>>>	4	5,9 Negative, < 0,3 g/L				
(46) Erba Lachema	CVP	>>>	11	16 Negative, < 0,3 g/L				
(49) BioVendor	CVP	>>>	1	1,5 Negative, < 0,3 g/L				
(60) Roche	CVP	>>>	4	5,9 Negative, < 0,3 g/L				
(63) Sysmex	CVP	>>>	4	5,9 Negative, < 0,3 g/L				
(109) Axis-Shield	CVP	>>>	1	1,5 Negative, < 0,3 g/L				
(162) Siemens (Atellica)	CVP	>>>	5	7,4 Negative, < 0,3 g/L				
(179) Siemens	CVP	>>>	5	7,4 Negative, < 0,3 g/L				
(201) Iris Diagnostics	CVP	>>>	1	1,5 Negative, < 0,3 g/L				
(202) Dirui	CVP	>>>	19	28 Negative, < 0,3 g/L				
(207) 77 Elektronika Kft.	CVP	>>>	6	8,8 Negative, < 0,3 g/L				
(999) another manufacturer	CVP	>>>	2	2,9 Negative, < 0,3 g/L				
Sample B		68				68	66	97
(5) Dialab	CVP	>>>	1	1,5 > 2,5 g/L				
(12) Beckman Coulter	CVP	>>>	3	4,4 0,51 - 2,5 g/L				
	CVP	>>>	1	1,5 > 2,5 g/L				
(18) Cormay	CVP	>>>	4	5,9 > 2,5 g/L				
(46) Erba Lachema	CVP	>>>	6	8,8 0,51 - 2,5 g/L				
	CVP	>>>	5	7,4 > 2,5 g/L				
(49) BioVendor	CVP	>>>	1	1,5 0,51 - 2,5 g/L				
(60) Roche	CVP	>>>	1	1,5 0,51 - 2,5 g/L				
	CVP	>>>	3	4,4 > 2,5 g/L				
(63) Sysmex	CVP	>>>	1	1,5 0,51 - 2,5 g/L				
	CVP	>>>	3	4,4 > 2,5 g/L				
(109) Axis-Shield	CVP	>>>	1	1,5 > 2,5 g/L				

Summary statistics - qualitative results**EQA round: DGP2/24 - Urine Strip Tests Analysis**

Deadline: 3.5.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success		
				Result		N _{eva}	N _{suc}	S _{rel} [%]
(162) Siemens (Atellica)	CVP >>>	3	4,4	0,51 - 2,5 g/L				
		2	2,9	> 2,5 g/L				
(179) Siemens	CVP >>>	1	1,5	0,51 - 2,5 g/L				
		4	5,9	> 2,5 g/L				
(201) Iris Diagnostics	CVP >>>	1	1,5	0,51 - 2,5 g/L				
(202) Dirui		2	2,9	0,3 - 0,5 g/L				
	CVP >>>	10	15	0,51 - 2,5 g/L				
		7	10	> 2,5 g/L				
(207) 77 Elektronika Kft.	CVP >>>	4	5,9	0,51 - 2,5 g/L				
		2	2,9	> 2,5 g/L				
(999) another manufacturer	CVP >>>	2	2,9	> 2,5 g/L				
(261) Glucose		70				70	65	93
Sample A		70				70	69	99
(5) Dialab	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L				
(12) Beckman Coulter	CVP >>>	4	5,7	Negative, normal, <= 2,8 mmol/L				
(18) Cormay	CVP >>>	4	5,7	Negative, normal, <= 2,8 mmol/L				
		1	1,4	11,1 - 28 mmol/L				
(46) Erba Lachema	CVP >>>	11	16	Negative, normal, <= 2,8 mmol/L				
(49) BioVendor	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L				
(60) Roche	CVP >>>	4	5,7	Negative, normal, <= 2,8 mmol/L				
(63) Sysmex	CVP >>>	4	5,7	Negative, normal, <= 2,8 mmol/L				
(109) Axis-Shield	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L				
(162) Siemens (Atellica)	CVP >>>	5	7,1	Negative, normal, <= 2,8 mmol/L				
(179) Siemens	CVP >>>	5	7,1	Negative, normal, <= 2,8 mmol/L				
(201) Iris Diagnostics	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L				
(202) Dirui	CVP >>>	19	27	Negative, normal, <= 2,8 mmol/L				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	Negative, normal, <= 2,8 mmol/L				
(999) another manufacturer	CVP >>>	3	4,3	Negative, normal, <= 2,8 mmol/L				
Sample B		70				70	66	94
(5) Dialab	CVP >>>	1	1,4	11,1 - 28 mmol/L				
(12) Beckman Coulter	CVP >>>	2	2,9	11,1 - 28 mmol/L				
	CVP >>>	2	2,9	> 28 mmol/L				
(18) Cormay		1	1,4	Negative, normal, <= 2,8 mmol/L				
		1	1,4	2,81 - 11 mmol/L				
	CVP >>>	3	4,3	> 28 mmol/L				
(46) Erba Lachema	CVP >>>	7	10	11,1 - 28 mmol/L				
	CVP >>>	4	5,7	> 28 mmol/L				
(49) BioVendor	CVP >>>	1	1,4	> 28 mmol/L				
(60) Roche	CVP >>>	4	5,7	> 28 mmol/L				
(63) Sysmex	CVP >>>	4	5,7	> 28 mmol/L				
(109) Axis-Shield	CVP >>>	1	1,4	11,1 - 28 mmol/L				
(162) Siemens (Atellica)	CVP >>>	5	7,1	11,1 - 28 mmol/L				
(179) Siemens	CVP >>>	4	5,7	11,1 - 28 mmol/L				
	CVP >>>	1	1,4	> 28 mmol/L				
(201) Iris Diagnostics	CVP >>>	1	1,4	> 28 mmol/L				
(202) Dirui		2	2,9	2,81 - 11 mmol/L				
	CVP >>>	13	19	11,1 - 28 mmol/L				
	CVP >>>	4	5,7	> 28 mmol/L				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	> 28 mmol/L				
(999) another manufacturer	CVP >>>	2	2,9	11,1 - 28 mmol/L				
	CVP >>>	1	1,4	> 28 mmol/L				
(275) hCG		3				3	3	100
Sample A		3				3	3	100
(17) Biogema	CVP >>>	1	33	Negative				
(179) Siemens	CVP >>>	1	33	Negative				
(202) Dirui	CVP >>>	1	33	Negative				
Sample B		3				3	3	100
(17) Biogema	CVP >>>	1	33	Positive				
(179) Siemens	CVP >>>	1	33	Positive				
(202) Dirui	CVP >>>	1	33	Positive				
(263) Ketone bodies		70				70	69	99
Sample A		70				70	69	99
(5) Dialab	CVP >>>	1	1,4	Negative, <= 0,5 mmol/L				
(12) Beckman Coulter	CVP >>>	4	5,7	Negative, <= 0,5 mmol/L				
(18) Cormay	CVP >>>	4	5,7	Negative, <= 0,5 mmol/L				
		1	1,4	1+ , 2+				
(46) Erba Lachema	CVP >>>	11	16	Negative, <= 0,5 mmol/L				
(49) BioVendor	CVP >>>	1	1,4	Negative, <= 0,5 mmol/L				
(60) Roche	CVP >>>	4	5,7	Negative, <= 0,5 mmol/L				

Summary statistics - qualitative results**EQA round: DGP2/24 - Urine Strip Tests Analysis**

Deadline: 3.5.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(63) Sysmex	CVP >>>	4	5,7	Negative, <= 0,5 mmol/L				
(109) Axis-Shield	CVP >>>	1	1,4	Negative, <= 0,5 mmol/L				
(162) Siemens (Atellica)	CVP >>>	5	7,1	Negative, <= 0,5 mmol/L				
(179) Siemens	CVP >>>	5	7,1	Negative, <= 0,5 mmol/L				
(201) Iris Diagnostics	CVP >>>	1	1,4	Negative, <= 0,5 mmol/L				
(202) Dirui	CVP >>>	19	27	Negative, <= 0,5 mmol/L				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	Negative, <= 0,5 mmol/L				
(999) another manufacturer	CVP >>>	3	4,3	Negative, <= 0,5 mmol/L				
Sample B		70				70	70	100
(5) Dialab	CVP >>>	1	1,4	3+, 4+				
(12) Beckman Coulter	CVP >>>	4	5,7	3+, 4+				
(18) Cormay	CVP >>>	5	7,1	3+, 4+				
(46) Erba Lachema	CVP >>>	6	8,6	1+, 2+				
	CVP >>>	5	7,1	3+, 4+				
(49) BioVendor	CVP >>>	1	1,4	3+, 4+				
(60) Roche	CVP >>>	1	1,4	1+, 2+				
	CVP >>>	3	4,3	3+, 4+				
(63) Sysmex	CVP >>>	4	5,7	3+, 4+				
(109) Axis-Shield	CVP >>>	1	1,4	3+, 4+				
(162) Siemens (Atellica)	CVP >>>	5	7,1	1+, 2+				
(179) Siemens	CVP >>>	2	2,9	1+, 2+				
	CVP >>>	3	4,3	3+, 4+				
(201) Iris Diagnostics	CVP >>>	1	1,4	1+, 2+				
(202) Dirui	CVP >>>	3	4,3	1+, 2+				
	CVP >>>	16	23	3+, 4+				
(207) 77 Elektronika Kft.	CVP >>>	1	1,4	1+, 2+				
	CVP >>>	5	7,1	3+, 4+				
(999) another manufacturer	CVP >>>	3	4,3	3+, 4+				
(272) Blood		70				70	66	94
Sample A		70				70	67	96
(5) Dialab	CVP >>>	1	1,4	Negative				
(12) Beckman Coulter	CVP >>>	4	5,7	Negative				
(18) Cormay	CVP >>>	4	5,7	Negative				
		1	1,4	5 - 12 ery/ μ L				
(46) Erba Lachema	CVP >>>	11	16	Negative				
(49) BioVendor	CVP >>>	1	1,4	Negative				
(60) Roche	CVP >>>	4	5,7	Negative				
(63) Sysmex	CVP >>>	4	5,7	Negative				
(109) Axis-Shield	CVP >>>	1	1,4	Negative				
(162) Siemens (Atellica)	CVP >>>	5	7,1	Negative				
(179) Siemens	CVP >>>	4	5,7	Negative				
		1	1,4	> 50 ery/ μ L				
(201) Iris Diagnostics	CVP >>>	1	1,4	Negative				
(202) Dirui	CVP >>>	19	27	Negative				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	Negative				
(999) another manufacturer	CVP >>>	2	2,9	Negative				
		1	1,4	5 - 12 ery/ μ L				
Sample B		70				70	69	99
(5) Dialab	CVP >>>	1	1,4	> 50 ery/ μ L				
(12) Beckman Coulter	CVP >>>	4	5,7	> 50 ery/ μ L				
(18) Cormay		1	1,4	Negative				
	CVP >>>	4	5,7	> 50 ery/ μ L				
(46) Erba Lachema	CVP >>>	11	16	> 50 ery/ μ L				
(49) BioVendor	CVP >>>	1	1,4	> 50 ery/ μ L				
(60) Roche		>	1	1,4	13 - 50 ery/ μ L			
	CVP >>>	3	4,3	> 50 ery/ μ L				
(63) Sysmex	CVP >>>	4	5,7	> 50 ery/ μ L				
(109) Axis-Shield	CVP >>>	1	1,4	> 50 ery/ μ L				
(162) Siemens (Atellica)		>	2	2,9	13 - 50 ery/ μ L			
	CVP >>>	3	4,3	> 50 ery/ μ L				
(179) Siemens	CVP >>>	5	7,1	> 50 ery/ μ L				
(201) Iris Diagnostics	CVP >>>	1	1,4	> 50 ery/ μ L				
(202) Dirui		>	1	1,4	13 - 50 ery/ μ L			
	CVP >>>	18	26	> 50 ery/ μ L				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	> 50 ery/ μ L				
(999) another manufacturer	CVP >>>	3	4,3	> 50 ery/ μ L				
(270) WBC		66				66	65	98
Sample A		66				66	65	98
(5) Dialab	CVP >>>	1	1,5	Negative				

Summary statistics - qualitative results**EQA round: DGP2/24 - Urine Strip Tests Analysis**

Setup: groups - R (strip manufacturer)

Deadline: 3.5.2024

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(12) Beckman Coulter	CVP >>>	4	6,1	Negative				
(18) Cormay	CVP >>>	5	7,6	Negative				
(46) Erba Lachema	CVP >>>	7	11	Negative				
		1	1,5	51 - 250 leu/µL				
(49) BioVendor	CVP >>>	1	1,5	Negative				
(60) Roche	CVP >>>	3	4,5	Negative				
(63) Sysmex	CVP >>>	4	6,1	Negative				
(109) Axis-Shield	CVP >>>	1	1,5	Negative				
(162) Siemens (Atellica)	CVP >>>	5	7,6	Negative				
(179) Siemens	CVP >>>	5	7,6	Negative				
(201) Iris Diagnostics	CVP >>>	1	1,5	Negative				
(202) Dirui	CVP >>>	19	29	Negative				
(207) 77 Elektronika Kft.	CVP >>>	6	9,1	Negative				
(999) another manufacturer	CVP >>>	3	4,5	Negative				
Sample B		66				66	66	100
(5) Dialab	CVP >>>	1	1,5	> 250 leu/µL				
(12) Beckman Coulter	CVP >>>	3	4,5	51 - 250 leu/µL				
	CVP >>>	1	1,5	> 250 leu/µL				
(18) Cormay	CVP >>>	3	4,5	51 - 250 leu/µL				
	CVP >>>	2	3	> 250 leu/µL				
(46) Erba Lachema	CVP >>>	1	1,5	51 - 250 leu/µL				
	CVP >>>	7	11	> 250 leu/µL				
(49) BioVendor	CVP >>>	1	1,5	> 250 leu/µL				
(60) Roche	CVP >>>	1	1,5	51 - 250 leu/µL				
	CVP >>>	2	3	> 250 leu/µL				
(63) Sysmex	CVP >>>	3	4,5	51 - 250 leu/µL				
	CVP >>>	1	1,5	> 250 leu/µL				
(109) Axis-Shield	CVP >>>	1	1,5	> 250 leu/µL				
(162) Siemens (Atellica)	CVP >>>	1	1,5	51 - 250 leu/µL				
	CVP >>>	4	6,1	> 250 leu/µL				
(179) Siemens	CVP >>>	1	1,5	51 - 250 leu/µL				
	CVP >>>	4	6,1	> 250 leu/µL				
(201) Iris Diagnostics	CVP >>>	1	1,5	51 - 250 leu/µL				
(202) Dirui	CVP >>>	5	7,6	51 - 250 leu/µL				
	CVP >>>	14	21	> 250 leu/µL				
(207) 77 Elektronika Kft.	CVP >>>	6	9,1	> 250 leu/µL				
(999) another manufacturer	CVP >>>	3	4,5	> 250 leu/µL				
(274) Nitrite		69				69	66	96
Sample A		69				69	67	97
(5) Dialab	CVP >>>	1	1,4	Negative				
(12) Beckman Coulter	CVP >>>	4	5,8	Negative				
(18) Cormay	CVP >>>	4	5,8	Negative				
		1	1,4	Positive				
(46) Erba Lachema	CVP >>>	9	13	Negative				
		1	1,4	Positive				
(49) BioVendor	CVP >>>	1	1,4	Negative				
(60) Roche	CVP >>>	4	5,8	Negative				
(63) Sysmex	CVP >>>	4	5,8	Negative				
(109) Axis-Shield	CVP >>>	1	1,4	Negative				
(162) Siemens (Atellica)	CVP >>>	5	7,2	Negative				
(179) Siemens	CVP >>>	5	7,2	Negative				
(201) Iris Diagnostics	CVP >>>	1	1,4	Negative				
(202) Dirui	CVP >>>	19	28	Negative				
(207) 77 Elektronika Kft.	CVP >>>	6	8,7	Negative				
(999) another manufacturer	CVP >>>	3	4,3	Negative				
Sample B		69				69	68	99
(5) Dialab	CVP >>>	1	1,4	Positive				
(12) Beckman Coulter	CVP >>>	4	5,8	Positive				
(18) Cormay		1	1,4	Not specified				
	CVP >>>	4	5,8	Positive				
(46) Erba Lachema	CVP >>>	10	14	Positive				
(49) BioVendor	CVP >>>	1	1,4	Positive				
(60) Roche	CVP >>>	4	5,8	Positive				
(63) Sysmex	CVP >>>	4	5,8	Positive				
(109) Axis-Shield	CVP >>>	1	1,4	Positive				
(162) Siemens (Atellica)	CVP >>>	5	7,2	Positive				
(179) Siemens	CVP >>>	5	7,2	Positive				
(201) Iris Diagnostics	CVP >>>	1	1,4	Positive				
(202) Dirui	CVP >>>	19	28	Positive				

Summary statistics - qualitative results**EQA round: DGP2/24 - Urine Strip Tests Analysis**

Deadline: 3.5.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success N _{eva} N _{suc} S _{rel} [%]
				Result		
(207) 77 Elektronika Kft.	CVP >>>	6	8,7	Positive		
(999) another manufacturer	CVP >>>	3	4,3	Positive		
(273) pH		70				70 67 96
Sample A		70				70 69 99
(5) Dialab	>	1	1,4	6,0		
(12) Beckman Coulter	CVP >>>	3	4,3	5,0		
	>	1	1,4	5,5		
(18) Cormay	CVP >>>	4	5,7	5,0		
	>	1	1,4	6,0		
(46) Erba Lachema	CVP >>>	8	11	5,0		
	>	2	2,9	5,5		
	>	1	1,4	6,0		
(49) BioVendor	CVP >>>	1	1,4	5,0		
(60) Roche	CVP >>>	3	4,3	5,0		
		1	1,4	7,0		
(63) Sysmex	CVP >>>	4	5,7	5,0		
(109) Axis-Shield	>	1	1,4	6,0		
(162) Siemens (Atellica)	>	5	7,1	5,5		
(179) Siemens	CVP >>>	1	1,4	5,0		
	>	2	2,9	5,5		
	>	2	2,9	6,0		
(201) Iris Diagnostics	>	1	1,4	5,5		
(202) Dirui	CVP >>>	13	19	5,0		
	>	5	7,1	5,5		
	>	1	1,4	6,0		
(207) 77 Elektronika Kft.	CVP >>>	5	7,1	5,0		
	>	1	1,4	5,5		
(999) another manufacturer	CVP >>>	1	1,4	5,0		
	>	2	2,9	6,0		
Sample B		70				70 68 97
(5) Dialab	>	1	1,4	7,5		
(12) Beckman Coulter	>	1	1,4	6,5		
	CVP >>>	3	4,3	7,0		
(18) Cormay	>	1	1,4	6,5		
	CVP >>>	1	1,4	7,0		
	>	3	4,3	7,5		
(46) Erba Lachema	>	1	1,4	6,0		
	CVP >>>	5	7,1	7,0		
	>	1	1,4	7,5		
	>	2	2,9	8,0		
	>	2	2,9	9,0		
(49) BioVendor	>	1	1,4	6,5		
(60) Roche	CVP >>>	1	1,4	7,0		
	>	3	4,3	8,0		
(63) Sysmex	>	4	5,7	7,5		
(109) Axis-Shield	>	1	1,4	7,5		
(162) Siemens (Atellica)	CVP >>>	5	7,1	7,0		
(179) Siemens	CVP >>>	3	4,3	7,0		
	>	2	2,9	7,5		
(201) Iris Diagnostics	CVP >>>	1	1,4	7,0		
(202) Dirui	>	1	1,4	6,5		
	CVP >>>	15	21	7,0		
	>	3	4,3	7,5		
(207) 77 Elektronika Kft.	>	5	7,1	6,5		
	CVP >>>	1	1,4	7,0		
(999) another manufacturer	CVP >>>	1	1,4	7,0		
	>	2	2,9	7,5		
(266) Urobilinogen		70				70 68 97
Sample A		70				70 70 100
(5) Dialab	CVP >>>	1	1,4	Negative, < 17 µmol/L		
(12) Beckman Coulter	CVP >>>	4	5,7	Negative, < 17 µmol/L		
(18) Cormay	CVP >>>	5	7,1	Negative, < 17 µmol/L		
(46) Erba Lachema	CVP >>>	11	16	Negative, < 17 µmol/L		
(49) BioVendor	CVP >>>	1	1,4	Negative, < 17 µmol/L		
(60) Roche	CVP >>>	4	5,7	Negative, < 17 µmol/L		
(63) Sysmex	CVP >>>	4	5,7	Negative, < 17 µmol/L		
(109) Axis-Shield	CVP >>>	1	1,4	Negative, < 17 µmol/L		
(162) Siemens (Atellica)	CVP >>>	5	7,1	Negative, < 17 µmol/L		
(179) Siemens	CVP >>>	5	7,1	Negative, < 17 µmol/L		

Summary statistics - qualitative results**EQA round: DGP2/24 - Urine Strip Tests Analysis**

Deadline: 3.5.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(201) Iris Diagnostics	CVP >>>	1	1,4	Negative, < 17 µmol/L				
(202) Dirui	CVP >>>	19	27	Negative, < 17 µmol/L				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	Negative, < 17 µmol/L				
(999) another manufacturer	CVP >>>	3	4,3	Negative, < 17 µmol/L				
Sample B		70				70	68	97
(5) Dialab	CVP >>>	1	1,4	1+, 2+				
(12) Beckman Coulter	CVP >>>	3	4,3	1+, 2+				
	CVP >>>	1	1,4	3+, 4+				
(18) Cormay	CVP >>>	1	1,4	1+, 2+				
	CVP >>>	4	5,7	3+, 4+				
(46) Erba Lachema	CVP >>>	1	1,4	1+, 2+				
	CVP >>>	10	14	3+, 4+				
(49) BioVendor	CVP >>>	1	1,4	3+, 4+				
(60) Roche	CVP >>>	4	5,7	3+, 4+				
(63) Sysmex	CVP >>>	3	4,3	1+, 2+				
	CVP >>>	1	1,4	3+, 4+				
(109) Axis-Shield	CVP >>>	1	1,4	3+, 4+				
(162) Siemens (Atellica)		1	1,4	Negative, < 17 µmol/L				
	CVP >>>	4	5,7	1+, 2+				
(179) Siemens		1	1,4	Negative, < 17 µmol/L				
	CVP >>>	1	1,4	1+, 2+				
	CVP >>>	3	4,3	3+, 4+				
(201) Iris Diagnostics	CVP >>>	1	1,4	3+, 4+				
(202) Dirui	CVP >>>	5	7,1	1+, 2+				
	CVP >>>	14	20	3+, 4+				
(207) 77 Elektronika Kft.	CVP >>>	6	8,6	3+, 4+				
(999) another manufacturer	CVP >>>	1	1,4	1+, 2+				
	CVP >>>	2	2,9	3+, 4+				