

EQA round: INRP2/24 - INR measurement on POCT

Deadline: 7.6.2024

Setup: all data

RoM = robust average

SD = standard deviation

CV = coefficient of variation

N_{tot} = total number of the resultsN_{out} = number of the results removed before calculation

AV = assigned value

CVP = consensus of all participants

U_{AV} = expanded uncertainty of the assigned value (k = 2)D_{max} = acceptable difference

LL = lower limit

UL = upper limit

N_{eva} = number of the results assessedN_{suc} = number of successful resultsS_{rel} = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]
							AV	U _{AV}	D _{max}	LL	UL			
(536) Prothrombin test (CoaguChek)	[-]				83							83	83	100
Sample A1												83	83	100
All results		3,06	0,17	5,6	83	0	CVP	3,06	0,046	20%	2,44	3,68	83	
Sample B1												83	83	100
All results		1,62	0,055	3,4	83	0	CVP	1,62	0,015	20%	1,29	1,95	83	