

## EQA round: AKS1/25 - Basic Clinical Chemistry - Serum

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

RoM = robust average	AV = assigned value	D <sub>max</sub> = acceptable difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	CVP = consensus of all participants	UL = upper limit
N <sub>tot</sub> = total number of the results	CVPG = consensus of the participants' groups	N <sub>eva</sub> = number of the results assessed
N <sub>out</sub> = number of the results removed before calculation	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	N <sub>suc</sub> = number of successful results
		S <sub>rel</sub> = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability					Traceability								
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>
<b>(22) α-amylase</b>	[μkat/L]				200							0						199	199	100
<b>Sample A</b>		5,93	0,39	6,5	200							0						199	199	100
(1) Abbott		6,29	0,11	1,8	45	0							CRV	5,98	0,16	15%	5,08	6,88		45
(46) Erba Lachema		6,13	0,38	6,2	7	0							CRV	5,98	0,16	15%	5,08	6,88		7
(58) Beckman Coulter (AU)		5,96	0,19	3,2	30	0							CRV	5,98	0,16	15%	5,08	6,88		30
(60) Roche		5,59	0,11	1,9	82	0							CRV	5,98	0,16	15%	5,08	6,88		82
(162) Siemens (Atellica)		6,34	0,18	2,8	23	0							CRV	5,98	0,16	15%	5,08	6,88		23
Other					13	0						0								12
							1x 149							1x 5, 2x 12, 1x 85, 1x 149, 4x 177, 4x 179						
<b>Sample B</b>		4,75	0,3	6,4	200							0						199	199	100
(1) Abbott		5,04	0,069	1,4	45	0							CRV	4,79	0,13	15%	4,07	5,51		45
(46) Erba Lachema		4,9	0,35	7,1	7	0							CRV	4,79	0,13	15%	4,07	5,51		7
(58) Beckman Coulter (AU)		4,77	0,14	3	30	0							CRV	4,79	0,13	15%	4,07	5,51		30
(60) Roche		4,49	0,077	1,7	82	0							CRV	4,79	0,13	15%	4,07	5,51		82
(162) Siemens (Atellica)		5,08	0,14	2,8	23	0							CRV	4,79	0,13	15%	4,07	5,51		23
Other					13	0						0								12
							1x 149							1x 5, 2x 12, 1x 85, 1x 149, 4x 177, 4x 179						
<b>(35) α-amylase pancreatic</b>	[μkat/L]				79							79	78	99						0
<b>Sample A</b>		4,97	0,18	3,6	79		CVP	4,97	0,05	18%	4,07	5,87	79	78	99					0
(1) Abbott		4,88	0,091	1,9	22	0							22							
(46) Erba Lachema		5,19	0,47	9,1	8	0							8							
(60) Roche		4,92	0,12	2,5	30	1							30							
(162) Siemens (Atellica)		5,67	0,19	3,3	10	0							10							
Other					9	0							9							
							1x 5, 3x 58, 3x 116, 2x 179													
<b>Sample B</b>		3,95	0,15	3,7	79		CVP	3,95	0,041	18%	3,23	4,67	79	79	100					0
(1) Abbott		3,89	0,079	2	22	0							22							
(46) Erba Lachema		4,12	0,39	9,4	8	0							8							
(60) Roche		3,9	0,09	2,3	30	0							30							
(162) Siemens (Atellica)		4,56	0,1	2,3	10	0							10							
Other					9	0							9							
							1x 5, 3x 58, 3x 116, 2x 179													
<b>(32) γ-globulin (elfo)</b>	[-]				52								52	50	96					0
<b>Sample A</b>		0,124	0,016	13	52		CVP	0,124	0,006	30%	0,086	0,162	52	50	96					0
(34) Sebia		0,121	0,016	13	41	0							41							
(213) Interlab Srl		0,137	0,013	9,7	10	0							10							
Other					1	0							1							
							1x 100													
<b>Sample B</b>		0,124	0,015	12	52		CVP	0,124	0,005	30%	0,086	0,162	52	50	96					0
(34) Sebia		0,121	0,015	12	41	0							41							

## EQA round: AKS1/25 - Basic Clinical Chemistry - Serum

Deadline: 24.1.2025

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability					N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	Traceability								
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL				AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	
(213) Interlab Srl		0,131	0,011	8,5	10	0						10											
Other					1	0						1											
							1x 100																
<b>(10) Albumin</b>	[g/L]				199							199	199	100							0		
<b>Sample A</b>		44,4	1,3	2,9	199		CVP	44,4	0,23	10%	39,9	48,9	199	199	100						0		
(1) Abbott		44,5	1,1	2,4	45	0							45										
(46) Erba Lachema		43,5	1,3	2,9	6	0							6										
(58) Beckman Coulter (AU)		43,4	1,3	3	30	0							30										
(60) Roche		44,7	1,2	2,7	82	0							82										
(162) Siemens (Atellica)		45,1	1,4	3,1	21	0							21										
(179) Siemens		44,5	0,74	1,7	6	0							6										
Other					9	0							9										
							2x 12, 1x 77, 1x 85, 1x 149, 1x 166, 3x 177																
<b>Sample B</b>		39,3	1,2	3,1	199		CVP	39,3	0,21	10%	35,3	43,3	199	199	100						0		
(1) Abbott		39	1	2,6	45	0							45										
(46) Erba Lachema		38,5	1,3	3,3	6	0							6										
(58) Beckman Coulter (AU)		38,3	1,2	3	30	0							30										
(60) Roche		40	1	2,5	82	0							82										
(162) Siemens (Atellica)		39,4	0,93	2,4	21	0							21										
(179) Siemens		39,3	1,1	2,8	6	0							6										
Other					9	0							9										
							2x 12, 1x 77, 1x 85, 1x 149, 1x 166, 3x 177																
<b>(31) Albumin (elpho)</b>	[-]				52							52	50	96							0		
<b>Sample A</b>		0,64	0,04	6,2	52		CVP	0,64	0,014	15%	0,544	0,736	52	50	96						0		
(34) Sebia		0,653	0,029	4,5	41	0							41										
(213) Interlab Srl		0,585	0,047	8	10	0							10										
Other					1	0							1										
							1x 100																
<b>Sample B</b>		0,64	0,035	5,4	52		CVP	0,64	0,012	15%	0,544	0,736	52	51	98						0		
(34) Sebia		0,651	0,024	3,7	41	0							41										
(213) Interlab Srl		0,596	0,042	7	10	0							10										
Other					1	0							1										
							1x 100																
<b>(21) ALP</b>	[μkat/L]				202							187	179	96							202	195	97
<b>Sample A</b>		0,784	0,056	7,1	202							187	179	96	CRV	0,828	0,035	18%	0,678	0,978	202	195	97
(1) Abbott		0,845	0,041	4,8	48	0	CVPG	0,845	0,014	14%	0,726	0,964	48								48		
(46) Erba Lachema		0,75	0,015	2	5	0	CVPG	0,75	0,042	14%	0,645	0,855	5								5		
(58) Beckman Coulter (AU)		0,725	0,064	8,9	30	0	CVPG	0,725	0,029	14%	0,623	0,827	30								30		
(60) Roche		0,771	0,03	3,8	81	0	CVPG	0,771	0,008	14%	0,663	0,879	81								81		
(162) Siemens (Atellica)		0,783	0,031	3,9	23	0	CVPG	0,783	0,016	14%	0,673	0,893	23								23		
Other					15	0							0								15		
							1x 5, 2x 12, 1x 85, 1x 149, 2x 158, 4x 177, 4x 179								1x 5, 2x 12, 1x 85, 1x 149, 2x 158, 4x 177, 4x 179								
<b>Sample B</b>		1,84	0,12	6,5	202							187	184	98	CRV	1,987	0,058	18%	1,62	2,35	202	201	100
(1) Abbott		1,95	0,054	2,8	48	0	CVPG	1,95	0,019	14%	1,67	2,23	48								48		
(46) Erba Lachema		1,86	0,059	3,2	5	0	CVPG	1,86	0,17	14%	1,59	2,13	5								5		
(58) Beckman Coulter (AU)		1,9	0,11	5,6	30	0	CVPG	1,9	0,048	14%	1,63	2,17	30								30		

## EQA round: AKS1/25 - Basic Clinical Chemistry - Serum

Deadline: 24.1.2025

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability					N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	Traceability					N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]		
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL				AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL					
(60) Roche		1,76	0,075	4,3	81	0	CVPG	1,76	0,02	14%	1,51	2,01	81								81			
(162) Siemens (Atellica)		1,79	0,053	2,9	23	0	CVPG	1,79	0,027	14%	1,53	2,05	23								23			
Other					15	0							0								15			
							1x 5, 2x 12, 1x 85, 1x 149, 2x 158, 4x 177, 4x 179								1x 5, 2x 12, 1x 85, 1x 149, 2x 158, 4x 177, 4x 179									
<b>(24) ALT</b>	[μkat/L]				207								0								207	197	95	
<b>Sample A</b>		1,72	0,092	5,3	207								0	CRV	1,702	0,045	15%	1,44	1,96		207	206	100	
(1) Abbott		1,76	0,07	4	48	0															48			
(46) Erba Lachema		1,85	0,082	4,4	5	0															5			
(58) Beckman Coulter (AU)		1,72	0,041	2,4	30	0															30			
(60) Roche		1,65	0,047	2,9	85	0															85			
(162) Siemens (Atellica)		1,81	0,028	1,6	24	0															24			
Other					15	0															15			
							2x 12, 1x 85, 1x 121, 1x 149, 4x 177, 2x 178, 4x 179																	
<b>Sample B</b>		0,871	0,051	5,8	207								0	CRV	0,827	0,022	15%	0,702	0,952		207	197	95	
(1) Abbott		0,882	0,044	5	48	0															48			
(46) Erba Lachema		0,94	0,045	4,7	5	0															5			
(58) Beckman Coulter (AU)		0,885	0,032	3,6	30	0															30			
(60) Roche		0,834	0,028	3,4	85	0															85			
(162) Siemens (Atellica)		0,92	0,022	2,4	24	0															24			
Other					15	0															15			
							2x 12, 1x 85, 1x 121, 1x 149, 4x 177, 2x 178, 4x 179																	
<b>(23) AST</b>	[μkat/L]				206								0								206	202	98	
<b>Sample A</b>		1,52	0,069	4,5	206								0	CRV	1,473	0,044	15%	1,25	1,7		206	202	98	
(1) Abbott		1,55	0,031	2	48	0															48			
(46) Erba Lachema		1,47	0,03	2	5	0															5			
(58) Beckman Coulter (AU)		1,55	0,082	5,2	30	0															30			
(60) Roche		1,47	0,063	4,3	85	0															85			
(162) Siemens (Atellica)		1,57	0,039	2,5	24	0															24			
Other					14	0															14			
							2x 12, 1x 85, 1x 149, 4x 177, 2x 178, 4x 179																	
<b>Sample B</b>		2,03	0,079	3,9	206								0	CRV	1,983	0,048	15%	1,68	2,29		206	203	99	
(1) Abbott		2,07	0,04	1,9	48	0															48			
(46) Erba Lachema		1,97	0,089	4,5	5	0															5			
(58) Beckman Coulter (AU)		2,08	0,1	4,9	30	0															30			
(60) Roche		1,98	0,078	3,9	85	0															85			
(162) Siemens (Atellica)		2,09	0,04	1,9	24	0															24			
Other					14	0															14			
							2x 12, 1x 85, 1x 149, 4x 177, 2x 178, 4x 179																	
<b>(13) Bilirubin total</b>	[μmol/L]				206								0								206	204	99	
<b>Sample A</b>		21,7	2,4	11	206								0	CRV	21,42	0,57	26%	15,8	27		206	204	99	
(1) Abbott		22,5	1	4,4	48	0															48			
(46) Erba Lachema		23,4	1,1	4,8	5	0															5			
(58) Beckman Coulter (AU)		24,3	0,71	2,9	28	0															28			
(60) Roche		19,5	0,66	3,4	85	0															85			
(162) Siemens (Atellica)		23,6	1,1	4,5	24	0															24			





## EQA round: AKS1/25 - Basic Clinical Chemistry - Serum

Deadline: 24.1.2025

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability					Traceability				
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	AV	U <sub>AV</sub>
Other					11	0	2x 12, 4x 46, 1x 77, 1x 149, 3x 177					11				
<b>(27) gamma-GT</b>	[μkat/L]				205							0				
<b>Sample A</b>		0,985	0,038	3,9	205							CRV 1,053 0,027 21% 0,831 1,28				
(1) Abbott		0,999	0,04	4	47	0						205 205 100				
(46) Erba Lachema		0,995	0,048	4,9	10	0						47				
(58) Beckman Coulter (AU)		0,963	0,026	2,7	29	0						10				
(60) Roche		0,975	0,026	2,6	83	0						29				
(162) Siemens (Atellica)		1,03	0,021	2	23	0						83				
Other					13	0						23				
												13				
												2x 12, 1x 85, 1x 149, 4x 177, 1x 178, 4x 179				
<b>Sample B</b>		2	0,075	3,8	205							CRV 2,052 0,053 21% 1,62 2,49				
(1) Abbott		1,99	0,079	4	47	0						205 204 100				
(46) Erba Lachema		2,05	0,084	4,1	10	0						47				
(58) Beckman Coulter (AU)		1,93	0,062	3,2	29	0						10				
(60) Roche		2	0,05	2,5	83	0						29				
(162) Siemens (Atellica)		2,08	0,041	2	23	0						83				
Other					13	0						23				
												13				
												2x 12, 1x 85, 1x 149, 4x 177, 1x 178, 4x 179				
<b>(16) Glucose</b>	[mmol/L]				209							0				
<b>Sample A</b>		4,47	0,11	2,4	209							CRV 4,656 0,047 8% 4,28 5,03				
(1) Abbott		4,41	0,087	2	46	0						209 206 99				
(46) Erba Lachema		4,53	0,045	0,98	9	0						46				
(58) Beckman Coulter (AU)		4,46	0,081	1,8	29	0						9				
(60) Roche		4,48	0,092	2	85	0						29				
(162) Siemens (Atellica)		4,49	0,11	2,4	24	0						85				
Other					16	0						24				
												16				
												1x 5, 2x 12, 1x 49, 1x 75, 1x 85, 1x 149, 1x 166, 4x 177, 4x 179				
<b>Sample B</b>		5,31	0,13	2,4	209							CRV 5,542 0,055 8% 5,09 5,99				
(1) Abbott		5,25	0,1	1,9	46	0						209 206 99				
(46) Erba Lachema		5,38	0,082	1,5	9	0						46				
(58) Beckman Coulter (AU)		5,31	0,13	2,5	29	0						9				
(60) Roche		5,32	0,12	2,2	85	0						29				
(162) Siemens (Atellica)		5,32	0,13	2,4	24	0						85				
Other					16	0						24				
												16				
												1x 5, 2x 12, 1x 49, 1x 75, 1x 85, 1x 149, 1x 166, 4x 177, 4x 179				
<b>(7) Magnesium</b>	[mmol/L]				191							0				
<b>Sample A</b>		1,51	0,044	2,9	191							CRV 1,499 0,029 15% 1,27 1,73				
(1) Abbott		1,53	0,053	3,4	47	0						191 189 99				
(58) Beckman Coulter (AU)		1,49	0,038	2,6	29	0						191 190 99				
(60) Roche		1,51	0,032	2,1	78	0						47				
(162) Siemens (Atellica)		1,52	0,039	2,5	20	0						29				
(179) Siemens		1,47	0,03	2	6	0						78				
Other					11	0						20				
												6				
												11				
												2x 12, 1x 46, 3x 77, 1x 85, 1x 149, 3x 177				













