

Cyklus EHK: AKS2/25 - Analyty krevního séra

Stop termín: 4.4.2025

Nastavení: skupiny - R (výrobce reagentů); minimální četnost skupin n = 5

Zkouška Vzorek Skupina	[jednotka]	RoM	SD	CV [%]	N _{tot}	N _{out}	Srovnatelnost					Návaznost						
							AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	AV	U _{AV}	D _{max}	LL
Ostatní					6	0	4x 58, 2x 178					6						
(32) γ-globulin (elfo)	[-]				85							85 84 99						
Vzorek A		0,123	0,013	11	85		CVP	0,123	0,004	30%	0,086	0,16	85	84	99	0		
(34) Sebia		0,124	0,013	11	63	0						63						
(213) Interlab Srl		0,117	0,013	11	20	0						20						
Ostatní					2	0						2						
							1x 63, 1x 100											
Vzorek B		0,124	0,014	12	85		CVP	0,124	0,004	30%	0,086	0,162	85	84	99	0		
(34) Sebia		0,123	0,015	12	63	0						63						
(213) Interlab Srl		0,124	0,013	11	20	0						20						
Ostatní					2	0						2						
							1x 63, 1x 100											
(10) Albumin	[g/L]				366							366 360 98						
Vzorek A		35,3	1,2	3,4	366		CVP	35,3	0,15	10%	31,7	38,9	366	360	98	0		
(1) Abbott		34,7	0,69	2	93	0						93						
(46) Erba Lachema		35,4	0,27	0,78	7	0						7						
(58) Beckman Coulter (AU)		34,3	1,1	3,2	49	0						49						
(60) Roche		36	0,85	2,3	153	0						153						
(162) Siemens (Atellica)		35,8	1,3	3,7	31	0						31						
(177) Mindray		34,2	0,86	2,5	11	0						11						
(179) Siemens		35	0,9	2,6	12	0						12						
Ostatní					10	0						10						
							2x 12, 1x 32, 1x 49, 2x 149, 1x 166, 3x 178											
Vzorek B		38,8	1,2	3,1	366		CVP	38,8	0,15	10%	34,9	42,7	366	364	99	0		
(1) Abbott		38,5	0,73	1,9	93	0						93						
(46) Erba Lachema		37,8	1,8	4,7	7	0						7						
(58) Beckman Coulter (AU)		37,7	1	2,7	49	0						49						
(60) Roche		39,3	0,9	2,3	153	0						153						
(162) Siemens (Atellica)		39,4	1,3	3,3	31	0						31						
(177) Mindray		37,7	1,2	3,1	11	0						11						
(179) Siemens		39	0,84	2,2	12	0						12						
Ostatní					10	0						10						
							2x 12, 1x 32, 1x 49, 2x 149, 1x 166, 3x 178											
(31) Albumin (elfo)	[-]				85							85 85 100						
Vzorek A		0,642	0,032	5	85		CVP	0,642	0,009	15%	0,545	0,739	85	85	100	0		
(34) Sebia		0,645	0,031	4,8	63	0						63						
(213) Interlab Srl		0,636	0,028	4,4	20	0						20						
Ostatní					2	0						2						
							1x 63, 1x 100											
Vzorek B		0,636	0,036	5,6	85		CVP	0,636	0,009	15%	0,54	0,732	85	85	100	0		
(34) Sebia		0,643	0,032	5	63	0						63						
(213) Interlab Srl		0,619	0,032	5,2	20	0						20						
Ostatní					2	0						2						
							1x 63, 1x 100											

Cyklus EHK: AKS2/25 - Analyty krevního séra

Stop termín: 4.4.2025

Nastavení: skupiny - R (výrobce reagentů); minimální četnost skupin n = 5

Zkouška		Srovnatelnost										Návaznost											
Vzorek	RoM	SD	CV	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]		
Skupina	[jednotka]		[%]																				
(21) ALP	[μkat/L]				383						376	366	97						383	313	82		
Vzorek A		1,34	0,11	8,3	383						376	368	98	CRV	1,239	0,035	18%	1,01	1,47	383	340	89	
(1) Abbott		1,43	0,052	3,6	100	0	CVPG	1,43	0,013	14%	1,22	1,64	100							100			
(46) Erba Lachema		1,39	0,14	9,7	10	0	CVPG	1,39	0,1	14%	1,19	1,59	10							10			
(58) Beckman Coulter (AU)		1,42	0,085	6	48	0	CVPG	1,42	0,03	14%	1,22	1,62	48							48			
(60) Roche		1,27	0,063	5	154	0	CVPG	1,27	0,012	14%	1,09	1,45	154							154			
(162) Siemens (Atellica)		1,27	0,044	3,5	34	0	CVPG	1,27	0,019	14%	1,09	1,45	34							34			
(177) Mindray		1,36	0,1	7,6	13	0	CVPG	1,36	0,071	14%	1,16	1,56	13							13			
(178) DiaSys		1,29	0,074	5,7	7	0	CVPG	1,29	0,072	14%	1,1	1,48	7							7			
(179) Siemens		1,27	0,073	5,8	10	0	CVPG	1,27	0,057	14%	1,09	1,45	10							10			
Ostatní					7	0							0							7			
											2x 12, 3x 149, 2x 158								2x 12, 3x 149, 2x 158				
Vzorek B		4,49	0,43	9,5	383						376	368	98	CRV	4,21	0,12	18%	3,45	4,97	383	332	87	
(1) Abbott		4,76	0,14	3	100	0	CVPG	4,76	0,035	14%	4,09	5,43	100							100			
(46) Erba Lachema		4,75	0,4	8,5	10	0	CVPG	4,75	0,31	14%	4,08	5,42	10							10			
(58) Beckman Coulter (AU)		5,03	0,24	4,7	48	0	CVPG	5,03	0,083	14%	4,32	5,74	48							48			
(60) Roche		4,18	0,23	5,4	154	0	CVPG	4,18	0,045	14%	3,59	4,77	154							154			
(162) Siemens (Atellica)		4,23	0,13	3	34	0	CVPG	4,23	0,053	14%	3,63	4,83	34							34			
(177) Mindray		4,82	0,3	6,2	13	0	CVPG	4,82	0,2	14%	4,14	5,5	13							13			
(178) DiaSys		4,35	0,25	5,8	7	0	CVPG	4,35	0,24	14%	3,74	4,96	7							7			
(179) Siemens		4,21	0,29	6,8	10	0	CVPG	4,21	0,22	14%	3,62	4,8	10							10			
Ostatní					7	0							0							7			
											2x 12, 3x 149, 2x 158								2x 12, 3x 149, 2x 158				
(24) ALT	[μkat/L]				389						0									389	373	96	
Vzorek A		0,953	0,055	5,8	389						0			CRV	0,923	0,023	15%	0,784	1,07	389	374	96	
(1) Abbott		0,962	0,046	4,8	102	0														102			
(46) Erba Lachema		0,996	0,053	5,3	10	0														10			
(58) Beckman Coulter (AU)		0,964	0,025	2,6	47	0														47			
(60) Roche		0,917	0,036	3,9	158	0														158			
(162) Siemens (Atellica)		1,01	0,038	3,8	35	0														35			
(177) Mindray		1,04	0,052	5	14	0														14			
(178) DiaSys		0,935	0,052	5,6	7	0														7			
(179) Siemens		1	0,048	4,8	11	0														11			
Ostatní					5	0														5			
																				2x 12, 1x 49, 2x 149			
Vzorek B		2,78	0,13	4,7	389						0			CRV	2,755	0,062	15%	2,34	3,17	389	384	99	
(1) Abbott		2,83	0,099	3,5	102	0														102			
(46) Erba Lachema		2,9	0,13	4,4	10	0														10			
(58) Beckman Coulter (AU)		2,8	0,069	2,5	47	0														47			
(60) Roche		2,69	0,079	2,9	158	0														158			
(162) Siemens (Atellica)		2,9	0,07	2,4	35	0														35			
(177) Mindray		2,92	0,13	4,5	14	0														14			
(178) DiaSys		2,76	0,12	4,3	7	0														7			
(179) Siemens		2,89	0,13	4,4	11	0														11			
Ostatní					5	0														5			
																				2x 12, 1x 49, 2x 149			

Cyklus EHK: AKS2/25 - Analyty krevního séra

Stop termín: 4.4.2025

Nastavení: skupiny - R (výrobce reagentů); minimální četnost skupin n = 5

Zkouška							Srovnatelnost					Návaznost										
Vzorek	RoM	SD	CV	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	
Skupina	[jednotka]		[%]																			
(23) AST	[μkat/L]				390						0									390	384	98
Vzorek A		1,05	0,048	4,6	390						0			CRV	1,044	0,023	15%	0,887	1,21	390	385	99
(1) Abbott		1,08	0,026	2,4	102	0														102		
(46) Erba Lachema		1,03	0,035	3,4	10	0														10		
(58) Beckman Coulter (AU)		1,06	0,064	6,1	48	0														48		
(60) Roche		1,03	0,036	3,5	158	0														158		
(162) Siemens (Atellica)		1,08	0,022	2	35	0														35		
(177) Mindray		1,03	0,046	4,5	14	0														14		
(178) DiaSys		0,99	0,074	7,5	7	0														7		
(179) Siemens		1,06	0,056	5,3	11	0														11		
Ostatní					5	0														5		
														2x 12, 1x 49, 2x 149								
Vzorek B		3,53	0,13	3,7	390						0			CRV	3,467	0,076	15%	2,94	3,99	390	387	99
(1) Abbott		3,58	0,075	2,1	102	0														102		
(46) Erba Lachema		3,41	0,14	4,1	10	0														10		
(58) Beckman Coulter (AU)		3,6	0,19	5,1	48	0														48		
(60) Roche		3,47	0,1	3	158	0														158		
(162) Siemens (Atellica)		3,65	0,064	1,8	35	0														35		
(177) Mindray		3,47	0,14	4	14	0														14		
(178) DiaSys		3,39	0,17	5	7	0														7		
(179) Siemens		3,57	0,28	7,8	11	0														11		
Ostatní					5	0														5		
														2x 12, 1x 49, 2x 149								
(13) Bilirubin celkový	[μmol/L]				388						0									388	382	98
Vzorek A		65,9	5,6	8,4	388						0			CRV	65,1	1,5	26%	48,1	82,1	388	383	99
(1) Abbott		68,4	3,2	4,7	103	0														103		
(46) Erba Lachema		73	4,1	5,6	8	0														8		
(58) Beckman Coulter (AU)		69,2	1,7	2,5	44	0														44		
(60) Roche		60,9	1,9	3,1	158	0														158		
(162) Siemens (Atellica)		71,4	1,9	2,7	34	0														34		
(177) Mindray		69,5	2,8	4	14	0														14		
(178) DiaSys		69,3	1,6	2,4	5	0														5		
(179) Siemens		70,1	5,6	8	13	0														13		
Ostatní					9	0														9		
														2x 12, 1x 49, 2x 70, 1x 115, 2x 149, 1x 158								
Vzorek B		32,3	3	9,4	388						0			CRV	32,1	0,74	26%	23,7	40,5	388	384	99
(1) Abbott		33,9	1,5	4,3	103	0														103		
(46) Erba Lachema		35,1	2,1	5,9	8	0														8		
(58) Beckman Coulter (AU)		34,7	0,91	2,6	44	0														44		
(60) Roche		29,4	0,94	3,2	158	0														158		
(162) Siemens (Atellica)		34,8	0,97	2,8	34	0														34		
(177) Mindray		32,4	1,3	4	14	0														14		
(178) DiaSys		34	1	3,1	5	0														5		
(179) Siemens		34,9	2,7	7,7	13	0														13		
Ostatní					9	0														9		
														2x 12, 1x 49, 2x 70, 1x 115, 2x 149, 1x 158								

Cyklus EHK: AKS2/25 - Analyty krevního séra

Stop termín: 4.4.2025

Nastavení: skupiny - R (výrobce reagentů); minimální četnost skupin n = 5

Zkouška		Srovnatelnost											Návaznost										
Vzorek	[jednotka]	RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	
Vzorek B		1	0,03	2,9	352		CVP	1	0,004	10%	0,9	1,1	352	342	97							0	
(1) Abbott		0,993	0,022	2,2	92	0							92										
(58) Beckman Coulter (AU)		1,01	0,027	2,7	47	0							47										
(60) Roche		1	0,027	2,7	143	0							143										
(162) Siemens (Atellica)		1,01	0,033	3,3	33	0							33										
(177) Mindray		0,939	0,04	4,3	11	0							11										
(179) Siemens		1,02	0,035	3,5	12	0							12										
Ostatní					14	0							14										

Cyklus EHK: AKS2/25 - Analyty krevního séra

Stop termín: 4.4.2025

Nastavení: skupiny - R (výrobce reagentů); minimální četnost skupin n = 5

Zkouška		Srovnatelnost					Návaznost															
Vzorek	RoM	SD	CV	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel} [%]	
Skupina	[jednotka]		[%]										[%]									[%]
(58) Beckman Coulter (AU)		59,9	2,3	3,9	44	0																44
(60) Roche		59,8	2,5	4,2	157	0																157
(162) Siemens (Atellica)		59,7	2,6	4,3	35	0																35
(177) Mindray		58,4	3,4	5,8	13	0																13
(179) Siemens		59	2,1	3,5	11	0																11
Ostatní					14	0																14
3x 5, 2x 12, 1x 49, 2x 149, 1x 166, 4x 178, 1x 999																						
Vzorek B		40,3	3,1	7,8	388						0			CRV	40,23	0,4	13%	35	45,5	388	354	91
(1) Abbott		40,8	4,2	10	104	0																104
(46) Erba Lachema		39	3,3	8,3	10	0																10
(58) Beckman Coulter (AU)		40,5	1,4	3,5	44	0																44
(60) Roche		40,3	2,6	6,3	157	0																157
(162) Siemens (Atellica)		40,7	2,7	6,7	35	0																35
(177) Mindray		38,4	3,4	8,8	13	0																13
(179) Siemens		39,9	2,8	6,9	11	0																11
Ostatní					14	0																14
3x 5, 2x 12, 1x 49, 2x 149, 1x 166, 4x 178, 1x 999																						
(17) Kyselina močová	[μmol/L]				385						0									385	377	98
Vzorek A		240	8,4	3,5	385						0			CRV	239,6	2,4	12%	210	269	385	378	98
(1) Abbott		238	4,8	2	102	0																102
(46) Erba Lachema		253	14	5,6	9	0																9
(58) Beckman Coulter (AU)		249	4	1,6	47	0																47
(60) Roche		235	5,7	2,4	156	0																156
(162) Siemens (Atellica)		249	5,1	2,1	33	0																33
(177) Mindray		239	8,6	3,6	14	0																14
(178) DiaSys		242	10	4,3	7	0																7
(179) Siemens		246	4,2	1,7	11	0																11
Ostatní					6	0																6
2x 12, 1x 49, 1x 115, 2x 149																						
Vzorek B		489	13	2,7	385						0			CRV	483,3	4,8	12%	425	542	385	380	99
(1) Abbott		486	7	1,4	102	0																102
(46) Erba Lachema		514	27	5,2	9	0																9
(58) Beckman Coulter (AU)		500	5,9	1,2	47	0																47
(60) Roche		482	12	2,4	156	0																156
(162) Siemens (Atellica)		504	9,4	1,9	33	0																33
(177) Mindray		485	14	2,9	14	0																14
(178) DiaSys		499	20	4	7	0																7
(179) Siemens		502	11	2,3	11	0																11
Ostatní					6	0																6
2x 12, 1x 49, 1x 115, 2x 149																						
(12) Laktát	[mmol/L]				162						162	156	96									0
Vzorek A		1,63	0,073	4,5	162		CVP	1,63	0,014	15%	1,38	1,88		162	158	98						0
(1) Abbott		1,7	0,03	1,8	23	0					23											
(58) Beckman Coulter (AU)		1,55	0,048	3,1	22	0					22											
(60) Roche		1,63	0,042	2,6	69	0					69											
(162) Siemens (Atellica)		1,66	0,056	3,4	25	0					25											

