

## EQA round: APLA1/25 - Antiphospholipid Antibodies

Deadline: 14.3.2025

Setup: all data

AV = assigned value

N<sub>tot</sub> = total number of the resultsN<sub>eva</sub> = number of the results assessed

CVP = consensus of all participants

N<sub>rel</sub> = relative number of the resultsN<sub>suc</sub> = number of successful results

&gt;&gt;&gt; ... expected result

S<sub>rel</sub> = relative success

&gt; ... acceptable result

± ... result not assessed

Test Sample Group	Frequency of the results				Success		
	AV	N <sub>tot</sub>	N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
<b>(465) anti-β2-glycoprotein 1 screening</b>		11			11	11	100
<b>Sample A</b>		11			11	11	100
All results	CVP >>>	11	100	Positive			
<b>Sample B</b>		11			11	11	100
All results	CVP >>>	11	100	Negative			
<b>(467) anti-β2-glycoprotein 1 IgG</b>		61			61	60	98
<b>Sample A</b>		61			61	60	98
All results		1	1,6	Negative			
	CVP >>>	60	98	Positive			
<b>Sample B</b>		61			61	61	100
All results	CVP >>>	61	100	Negative			
<b>(468) anti-β2-glycoprotein 1 IgM</b>		60			60	60	100
<b>Sample A</b>		60			60	60	100
All results	CVP >>>	60	100	Negative			
<b>Sample B</b>		60			60	60	100
All results	CVP >>>	60	100	Negative			
<b>(461) anti-cardiolipin screening</b>		13			13	13	100
<b>Sample A</b>		13			13	13	100
All results	CVP >>>	13	100	Positive			
<b>Sample B</b>		13			13	13	100
All results	CVP >>>	13	100	Negative			
<b>(463) anti-cardiolipin IgG</b>		64			64	63	98
<b>Sample A</b>		64			64	63	98
All results		1	1,6	Negative			
	CVP >>>	63	98	Positive			
<b>Sample B</b>		64			64	64	100
All results	CVP >>>	64	100	Negative			
<b>(464) anti-cardiolipin IgM</b>		63			63	63	100
<b>Sample A</b>		63			63	63	100
All results	CVP >>>	63	100	Negative			
<b>Sample B</b>		63			63	63	100
All results	CVP >>>	63	100	Negative			