

EQA round: HKG1/25 - Haemocoagulation Tests

Deadline: 28.3.2025

Setup: minimal size of the groups n = 5

RoM = robust average	AV = assigned value	D _{max} = acceptable difference
SD = standard deviation	CVP = consensus of all participants	LL = lower limit
CV = coefficient of variation	CVPG = consensus of the participants' groups	UL = upper limit
N _{tot} = total number of the results	U _{AV} = expanded uncertainty of the assigned value (k = 2)	N _{eva} = number of the results assessed
N _{out} = number of the results removed before calculation		N _{suc} = number of successful results
		S _{rel} = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]
							AV	U _{AV}	D _{max}	LL	UL			
(177) Antithrombin	[%]				128							128	127	99
Sample A												128	128	100
All results		92,9	4	4,3	128	0	CVP	92,9	0,87	18%	76,1	110	128	
Sample B												128	127	99
All results		41,4	3	7,3	128	0	CVP	41,4	0,65	25%	31	51,8	128	
(170) APTT - ratio	[-]				210							201	197	98
Sample A		1,01	0,11	11	210							201	199	99
(240) Siemens (Dade) Actin FS		0,917	0,037	4,1	59	0	CVPG	0,917	0,012	20%	0,733	1,11	59	
(241) Siemens (Dade) Actin FSL		0,995	0,057	5,8	15	0	CVPG	0,995	0,036	20%	0,796	1,2	15	
(242) Siemens (Dade) Pathromtin SL		1,11	0,045	4,1	64	0	CVPG	1,11	0,014	20%	0,888	1,34	64	
(246) Mindray APTT Reagent		2,03	0,12	6	13	0	CVPG	2,03	0,083	20%	1,62	2,44	13	
(253) IL APTT-SP		0,925	0,027	2,9	21	0	CVPG	0,925	0,014	20%	0,74	1,11	21	
(256) IL SynthASil		1,04	0,03	2,9	9	0	CVPG	1,04	0,037	20%	0,832	1,25	9	
(260) Stago PTT Automate		1	0,037	3,6	20	0	CVPG	1	0,02	20%	0,8	1,2	20	
Other					9	0							0	
								1x 257, 1x 259, 3x 266, 2x 269, 2x 999						
Sample B		1,67	0,3	18	210							201	199	99
(240) Siemens (Dade) Actin FS		1,47	0,14	9,3	59	0	CVPG	1,47	0,044	20%	1,17	1,77	59	
(241) Siemens (Dade) Actin FSL		1,63	0,12	7,7	15	0	CVPG	1,63	0,079	20%	1,3	1,96	15	
(242) Siemens (Dade) Pathromtin SL		1,92	0,12	6,2	64	0	CVPG	1,92	0,037	20%	1,53	2,31	64	
(246) Mindray APTT Reagent		2,84	0,19	6,7	13	0	CVPG	2,84	0,13	20%	2,27	3,41	13	
(253) IL APTT-SP		1,34	0,027	2	21	0	CVPG	1,34	0,014	20%	1,07	1,61	21	
(256) IL SynthASil		1,41	0,059	4,2	9	0	CVPG	1,41	0,073	20%	1,12	1,7	9	
(260) Stago PTT Automate		1,66	0,083	5	20	0	CVPG	1,66	0,046	20%	1,32	2	20	
Other					9	0							0	
								1x 257, 1x 259, 3x 266, 2x 269, 2x 999						
(171) Fibrinogen	[g/L]				168							168	167	99
Sample A												168	168	100
All results		2,71	0,19	6,9	168	0	CVP	2,71	0,035	25%	2,03	3,39	168	
Sample B												168	167	99
All results		1,09	0,11	9,9	168	0	CVP	1,09	0,02	25%	0,817	1,37	168	
(172) Prothrombin test (INR)	[-]				216							216	211	98
Sample A		1,01	0,051	5,1	216	1	CVP	1,01	0,009	20%	0,808	1,22	216	215
All results		1,63	0,15	9,2	216							216	212	98
Sample B		1,6	0,12	7,3	183	0	CVP	1,6	0,021	20%	1,28	1,92	183	
All results (without individual groups)		1,95	0,16	8,4	33	0	CVPG	1,95	0,07	20%	1,56	2,34	33	
(3) Coagulometr (trom. rabbit br.)														
(179) Prothrombin test (ratio)	[-]				214							214	208	97
Sample A		1	0,05	5	214	0	CVP	1	0,008	20%	0,8	1,2	214	213
All results		1,59	0,14	8,8	214							214	209	98
Sample B		1,56	0,11	7,3	181	0	CVP	1,56	0,021	20%	1,24	1,88	181	
All results (without individual groups)		1,84	0,15	8,1	33	0	CVPG	1,84	0,063	20%	1,47	2,21	33	
(3) Coagulometr (trom. rabbit br.)														
(173) Thrombin time (time)	[s]				120							111	107	96
Sample A		15,6	0,85	5,5	120							111	111	100
(3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago		16	0,9	5,6	12	0	CVPG	16	0,64	15%	13,6	18,4	12	
(3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		18,7	0,52	2,8	5	0	CVPG	18,7	1,5	15%	15,8	21,6	5	
(5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		15,5	0,61	4	69	0	CVPG	15,5	0,18	15%	13,1	17,9	69	
(6) Thrombin units: 3 IU/mL (NIH); (38) Werfen (IL)		16	0,57	3,6	16	0	CVPG	16	0,35	15%	13,6	18,4	16	
(8) Počet trombin. jednotek: 8 IU/mL (NIH); (177) Mindray		15,1	1,5	9,8	9	0	CVPG	15,1	1,8	15%	12,8	17,4	9	
Other					9	0							0	
								2x 1/149, 1x 1/177, 2x 2/193, 2x 5/179, 1x 6/46, 1x 6/177						
Sample B		18,1	4,2	23	120							111	107	96
(3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago		17,4	0,46	2,7	12	0	CVPG	17,4	0,33	15%	14,7	20,1	12	
(3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		21,3	0,52	2,4	5	0	CVPG	21,3	1,5	15%	18,1	24,5	5	
(5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		15,6	0,78	5	69	1	CVPG	15,6	0,23	15%	13,2	18	69	
(6) Thrombin units: 3 IU/mL (NIH); (38) Werfen (IL)		26	1,2	4,6	16	0	CVPG	26	0,73	15%	22,1	29,9	16	

EQA round: HKG1/25 - Haemocoagulation Tests

Deadline: 28.3.2025

Setup: minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]	
							AV	U _{AV}	D _{max}	LL	UL				
(8) Počet trombin. jednotek: 8 IU/mL (NIH); (177)		23,7	1,8	7,6	9	0	CVPG	23,7	2,2	15%	20,1	27,3	9		
Mindray															
Other					9	0							0		
							2x 1/149, 1x 1/177, 2x 2/193, 2x 5/179, 1x 6/46, 1x 6/177								
(174) Thrombin time (ratio)	[-]				101								94	92	98
Sample A		0,929	0,098	11	101								94	94	100
(3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago		0,923	0,019	2,1	12	0	CVPG	0,923	0,014	20%	0,738	1,11	12		
(3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		1,03	0,1	10	5	0	CVPG	1,03	0,29	20%	0,824	1,24	5		
(5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		0,88	0,056	6,4	53	0	CVPG	0,88	0,019	20%	0,704	1,06	53		
(6) Thrombin units: 3 IU/mL (NIH); (38) Werfen (IL)		1,09	0,038	3,5	15	0	CVPG	1,09	0,024	20%	0,872	1,31	15		
(8) Počet trombin. jednotek: 8 IU/mL (NIH); (177)		0,98	0,015	1,5	9	0	CVPG	0,98	0,018	20%	0,784	1,18	9		
Mindray															
Other					7	0							0		
							1x 1/149, 2x 2/193, 2x 5/179, 1x 6/46, 1x 6/51								
Sample B		1,11	0,35	31	101								94	92	98
(3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago		0,993	0,049	5	12	0	CVPG	0,993	0,035	20%	0,794	1,2	12		
(3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		1,18	0,12	10	5	0	CVPG	1,18	0,34	20%	0,944	1,42	5		
(5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension)		0,898	0,068	7,5	53	1	CVPG	0,898	0,023	20%	0,718	1,08	53		
(6) Thrombin units: 3 IU/mL (NIH); (38) Werfen (IL)		1,77	0,086	4,9	15	0	CVPG	1,77	0,055	20%	1,41	2,13	15		
(8) Počet trombin. jednotek: 8 IU/mL (NIH); (177)		1,59	0,03	1,9	9	0	CVPG	1,59	0,037	20%	1,27	1,91	9		
Mindray															
Other					7	0							0		
							1x 1/149, 2x 2/193, 2x 5/179, 1x 6/46, 1x 6/51								