

## EQA round: DGP2/26 - Urine Strip Tests Analysis

Deadline: 1.5.2026

Setup: groups - R (strip manufacturer)

AV = assigned value

N<sub>tot</sub> = total number of the resultsN<sub>eva</sub> = number of the results assessed

CVP = consensus of all participants

N<sub>rel</sub> = relative number of the resultsN<sub>suc</sub> = number of successful results

CVPG = consensus of the participants' groups

S<sub>rel</sub> = relative success

&gt;&gt;&gt; ... expected result

&gt; ... acceptable result

± ... result not assessed

Test Sample Group	AV	N <sub>tot</sub>	Frequency of the results			Success		
			N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	
<b>(268) Bilirubin</b>		69				69	68	99
<b>Sample A</b>		69				69	68	99
(5) Dialab	CVP >>>	1	1,4	Negative				
(12) Beckman Coulter	CVP >>>	3	4,3	Negative				
(18) Cormay	CVP >>>	6	8,7	Negative				
(46) Erba Lachema	CVP >>>	8	12	Negative				
		1	1,4	1+, 2+				
(49) BioVendor	CVP >>>	1	1,4	Negative				
(60) Roche	CVP >>>	3	4,3	Negative				
(63) Sysmex	CVP >>>	7	10	Negative				
(109) Axis-Shield	CVP >>>	1	1,4	Negative				
(162) Siemens (Atellica)	CVP >>>	2	2,9	Negative				
(177) Mindray	CVP >>>	1	1,4	Negative				
(179) Siemens	CVP >>>	10	14	Negative				
(197) Analyticon	CVP >>>	1	1,4	Negative				
(202) Dirui	CVP >>>	19	28	Negative				
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	Negative				
(999) another manufacturer	CVP >>>	3	4,3	Negative				
<b>Sample B</b>		69				69	69	100
(5) Dialab	CVP >>>	1	1,4	3+, 4+				
(12) Beckman Coulter	>	3	4,3	1+, 2+				
(18) Cormay	CVP >>>	6	8,7	3+, 4+				
(46) Erba Lachema	>	1	1,4	1+, 2+				
	CVP >>>	8	12	3+, 4+				
(49) BioVendor	CVP >>>	1	1,4	3+, 4+				
(60) Roche	>	1	1,4	1+, 2+				
	CVP >>>	2	2,9	3+, 4+				
(63) Sysmex	CVP >>>	7	10	3+, 4+				
(109) Axis-Shield	CVP >>>	1	1,4	3+, 4+				
(162) Siemens (Atellica)	CVP >>>	2	2,9	3+, 4+				
(177) Mindray	CVP >>>	1	1,4	3+, 4+				
(179) Siemens	CVP >>>	10	14	3+, 4+				
(197) Analyticon	CVP >>>	1	1,4	3+, 4+				
(202) Dirui	CVP >>>	19	28	3+, 4+				
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	3+, 4+				
(999) another manufacturer	CVP >>>	3	4,3	3+, 4+				
<b>(264) Total protein</b>		69				69	69	100
<b>Sample A</b>		69				69	69	100
(5) Dialab	CVP >>>	1	1,4	Negative, < 0,3 g/L				
(12) Beckman Coulter	CVP >>>	3	4,3	Negative, < 0,3 g/L				
(18) Cormay	CVP >>>	5	7,2	Negative, < 0,3 g/L				
(46) Erba Lachema	CVP >>>	10	14	Negative, < 0,3 g/L				
(49) BioVendor	CVP >>>	1	1,4	Negative, < 0,3 g/L				
(60) Roche	CVP >>>	3	4,3	Negative, < 0,3 g/L				
(63) Sysmex	CVP >>>	7	10	Negative, < 0,3 g/L				
(109) Axis-Shield	CVP >>>	1	1,4	Negative, < 0,3 g/L				
(162) Siemens (Atellica)	CVP >>>	2	2,9	Negative, < 0,3 g/L				
(177) Mindray	CVP >>>	1	1,4	Negative, < 0,3 g/L				
(179) Siemens	CVP >>>	10	14	Negative, < 0,3 g/L				
(197) Analyticon	CVP >>>	1	1,4	Negative, < 0,3 g/L				
(202) Dirui	CVP >>>	19	28	Negative, < 0,3 g/L				
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	Negative, < 0,3 g/L				
(999) another manufacturer	CVP >>>	3	4,3	Negative, < 0,3 g/L				
<b>Sample B</b>		69				69	69	100
(5) Dialab	CVP >>>	1	1,4	> 2,5 g/L				
(12) Beckman Coulter	CVP >>>	3	4,3	> 2,5 g/L				
(18) Cormay	CVP >>>	5	7,2	> 2,5 g/L				
(46) Erba Lachema	>	2	2,9	0,51 - 2,5 g/L				
	CVP >>>	8	12	> 2,5 g/L				
(49) BioVendor	>	1	1,4	0,51 - 2,5 g/L				
(60) Roche	CVP >>>	3	4,3	> 2,5 g/L				
(63) Sysmex	CVP >>>	7	10	> 2,5 g/L				
(109) Axis-Shield	CVP >>>	1	1,4	> 2,5 g/L				

## EQA round: DGP2/26 - Urine Strip Tests Analysis

Deadline: 1.5.2026

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N <sub>tot</sub>	Frequency of the results		Success		
			N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
(162) Siemens (Atellica)	CVP >>>	2	2,9	> 2,5 g/L			
(177) Mindray	CVP >>>	1	1,4	> 2,5 g/L			
(179) Siemens	>	1	1,4	0,51 - 2,5 g/L			
	CVP >>>	9	13	> 2,5 g/L			
(197) Analyticon	>	1	1,4	0,51 - 2,5 g/L			
(202) Dirui	>	4	5,8	0,51 - 2,5 g/L			
	CVP >>>	15	22	> 2,5 g/L			
(207) 77 Elektronika Kft.	>	2	2,9	0,51 - 2,5 g/L			
(999) another manufacturer	CVP >>>	3	4,3	> 2,5 g/L			
<b>(261) Glucose</b>		70			70	69	99
<b>Sample A</b>		70			70	70	100
(5) Dialab	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L			
(12) Beckman Coulter	CVP >>>	3	4,3	Negative, normal, <= 2,8 mmol/L			
(18) Cormay	CVP >>>	6	8,6	Negative, normal, <= 2,8 mmol/L			
(46) Erba Lachema	CVP >>>	9	13	Negative, normal, <= 2,8 mmol/L			
(49) BioVendor	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L			
(60) Roche	CVP >>>	4	5,7	Negative, normal, <= 2,8 mmol/L			
(63) Sysmex	CVP >>>	7	10	Negative, normal, <= 2,8 mmol/L			
(109) Axis-Shield	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L			
(162) Siemens (Atellica)	CVP >>>	2	2,9	Negative, normal, <= 2,8 mmol/L			
(177) Mindray	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L			
(179) Siemens	CVP >>>	10	14	Negative, normal, <= 2,8 mmol/L			
(197) Analyticon	CVP >>>	1	1,4	Negative, normal, <= 2,8 mmol/L			
(202) Dirui	CVP >>>	19	27	Negative, normal, <= 2,8 mmol/L			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	Negative, normal, <= 2,8 mmol/L			
(999) another manufacturer	CVP >>>	3	4,3	Negative, normal, <= 2,8 mmol/L			
<b>Sample B</b>		70			70	69	99
(5) Dialab	CVP >>>	1	1,4	11,1 - 28 mmol/L			
(12) Beckman Coulter	CVP >>>	2	2,9	11,1 - 28 mmol/L			
	>	1	1,4	> 28 mmol/L			
(18) Cormay	CVP >>>	5	7,1	11,1 - 28 mmol/L			
	>	1	1,4	> 28 mmol/L			
(46) Erba Lachema	CVP >>>	5	7,1	11,1 - 28 mmol/L			
	>	4	5,7	> 28 mmol/L			
(49) BioVendor	>	1	1,4	> 28 mmol/L			
(60) Roche	>	1	1,4	2,81 - 11 mmol/L			
	>	3	4,3	> 28 mmol/L			
(63) Sysmex	CVP >>>	2	2,9	11,1 - 28 mmol/L			
	>	5	7,1	> 28 mmol/L			
(109) Axis-Shield	CVP >>>	1	1,4	11,1 - 28 mmol/L			
(162) Siemens (Atellica)	CVP >>>	1	1,4	11,1 - 28 mmol/L			
	>	1	1,4	> 28 mmol/L			
(177) Mindray	CVP >>>	1	1,4	11,1 - 28 mmol/L			
(179) Siemens	>	1	1,4	2,81 - 11 mmol/L			
	CVP >>>	7	10	11,1 - 28 mmol/L			
	>	2	2,9	> 28 mmol/L			
(197) Analyticon	>	1	1,4	> 28 mmol/L			
(202) Dirui	>	8	11	2,81 - 11 mmol/L			
	CVP >>>	9	13	11,1 - 28 mmol/L			
	>	2	2,9	> 28 mmol/L			
(207) 77 Elektronika Kft.	>	2	2,9	> 28 mmol/L			
(999) another manufacturer		1	1,4	Negative, normal, <= 2,8 mmol/L			
	CVP >>>	1	1,4	11,1 - 28 mmol/L			
	>	1	1,4	> 28 mmol/L			
<b>(275) hCG</b>		2			2	2	100
<b>Sample A</b>		2			2	2	100
(17) Biogema	CVP >>>	1	50	Negative			
(179) Siemens	CVP >>>	1	50	Negative			
<b>Sample B</b>		2			2	2	100
(17) Biogema	CVP >>>	1	50	Positive			
(179) Siemens	CVP >>>	1	50	Positive			
<b>(263) Ketone bodies</b>		70			70	70	100
<b>Sample A</b>		70			70	70	100
(5) Dialab	CVP >>>	1	1,4	Negative, <= 0,5 mmol/L			
(12) Beckman Coulter	CVP >>>	3	4,3	Negative, <= 0,5 mmol/L			
(18) Cormay	CVP >>>	6	8,6	Negative, <= 0,5 mmol/L			
(46) Erba Lachema	CVP >>>	10	14	Negative, <= 0,5 mmol/L			
(49) BioVendor	CVP >>>	1	1,4	Negative, <= 0,5 mmol/L			
(60) Roche	CVP >>>	3	4,3	Negative, <= 0,5 mmol/L			

## EQA round: DGP2/26 - Urine Strip Tests Analysis

Deadline: 1.5.2026

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N <sub>tot</sub>	Frequency of the results		Result	Success		
			N <sub>rel</sub> [%]			N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
(63) Sysmex	CVP >>>	7	10		Negative, <= 0,5 mmol/L			
(109) Axis-Shield	CVP >>>	1	1,4		Negative, <= 0,5 mmol/L			
(162) Siemens (Atellica)	CVP >>>	2	2,9		Negative, <= 0,5 mmol/L			
(177) Mindray	CVP >>>	1	1,4		Negative, <= 0,5 mmol/L			
(179) Siemens	CVP >>>	10	14		Negative, <= 0,5 mmol/L			
(197) Analyticon	CVP >>>	1	1,4		Negative, <= 0,5 mmol/L			
(202) Dirui	CVP >>>	19	27		Negative, <= 0,5 mmol/L			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9		Negative, <= 0,5 mmol/L			
(999) another manufacturer	CVP >>>	3	4,3		Negative, <= 0,5 mmol/L			
<b>Sample B</b>			70			70	70	100
(5) Dialab	>	1	1,4		1+, 2+			
(12) Beckman Coulter	CVP >>>	3	4,3		3+, 4+			
(18) Cormay	>	2	2,9		1+, 2+			
	CVP >>>	4	5,7		3+, 4+			
(46) Erba Lachema	>	3	4,3		1+, 2+			
	CVP >>>	7	10		3+, 4+			
(49) BioVendor	CVP >>>	1	1,4		3+, 4+			
(60) Roche	CVP >>>	3	4,3		3+, 4+			
(63) Sysmex	>	5	7,1		1+, 2+			
	CVP >>>	2	2,9		3+, 4+			
(109) Axis-Shield	CVP >>>	1	1,4		3+, 4+			
(162) Siemens (Atellica)	>	2	2,9		1+, 2+			
(177) Mindray	>	1	1,4		1+, 2+			
(179) Siemens	>	5	7,1		1+, 2+			
	CVP >>>	5	7,1		3+, 4+			
(197) Analyticon	>	1	1,4		1+, 2+			
(202) Dirui	>	2	2,9		1+, 2+			
	CVP >>>	17	24		3+, 4+			
(207) 77 Elektronika Kft.	>	1	1,4		1+, 2+			
	CVP >>>	1	1,4		3+, 4+			
(999) another manufacturer	>	1	1,4		1+, 2+			
	CVP >>>	2	2,9		3+, 4+			
<b>(272) Blood</b>			70			70	70	100
<b>Sample A</b>			70			70	70	100
(5) Dialab	CVP >>>	1	1,4		Negative			
(12) Beckman Coulter	CVP >>>	3	4,3		Negative			
(18) Cormay	CVP >>>	6	8,6		Negative			
(46) Erba Lachema	CVP >>>	10	14		Negative			
(49) BioVendor	CVP >>>	1	1,4		Negative			
(60) Roche	CVP >>>	3	4,3		Negative			
(63) Sysmex	CVP >>>	7	10		Negative			
(109) Axis-Shield	CVP >>>	1	1,4		Negative			
(162) Siemens (Atellica)	CVP >>>	2	2,9		Negative			
(177) Mindray	CVP >>>	1	1,4		Negative			
(179) Siemens	CVP >>>	10	14		Negative			
(197) Analyticon	CVP >>>	1	1,4		Negative			
(202) Dirui	CVP >>>	19	27		Negative			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9		Negative			
(999) another manufacturer	CVP >>>	3	4,3		Negative			
<b>Sample B</b>			70			70	70	100
(5) Dialab	CVP >>>	1	1,4		> 50 ery/ $\mu$ L			
(12) Beckman Coulter	CVP >>>	3	4,3		> 50 ery/ $\mu$ L			
(18) Cormay	>	1	1,4		13 - 50 ery/ $\mu$ L			
	CVP >>>	5	7,1		> 50 ery/ $\mu$ L			
(46) Erba Lachema	CVP >>>	10	14		> 50 ery/ $\mu$ L			
(49) BioVendor	CVP >>>	1	1,4		> 50 ery/ $\mu$ L			
(60) Roche	CVP >>>	3	4,3		> 50 ery/ $\mu$ L			
(63) Sysmex	CVP >>>	7	10		> 50 ery/ $\mu$ L			
(109) Axis-Shield	CVP >>>	1	1,4		> 50 ery/ $\mu$ L			
(162) Siemens (Atellica)	CVP >>>	2	2,9		> 50 ery/ $\mu$ L			
(177) Mindray	CVP >>>	1	1,4		> 50 ery/ $\mu$ L			
(179) Siemens	>	2	2,9		13 - 50 ery/ $\mu$ L			
	CVP >>>	8	11		> 50 ery/ $\mu$ L			
(197) Analyticon	CVP >>>	1	1,4		> 50 ery/ $\mu$ L			
(202) Dirui	>	5	7,1		13 - 50 ery/ $\mu$ L			
	CVP >>>	14	20		> 50 ery/ $\mu$ L			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9		> 50 ery/ $\mu$ L			
(999) another manufacturer	CVP >>>	3	4,3		> 50 ery/ $\mu$ L			

## EQA round: DGP2/26 - Urine Strip Tests Analysis

Deadline: 1.5.2026

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N <sub>tot</sub>	Frequency of the results			Success		
			N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]	
<b>(270) WBC</b>		66				63	59	94
<b>Sample A</b>		66				66	64	97
(5) Dialab	CVP >>>	1	1,5	Negative				
(12) Beckman Coulter	CVP >>>	3	4,5	Negative				
(18) Cormay	CVP >>>	6	9,1	Negative				
(46) Erba Lachema	CVP >>>	5	7,6	Negative				
		2	3	10 - 50 leu/μL				
(49) BioVendor	CVP >>>	1	1,5	Negative				
(60) Roche	CVP >>>	2	3	Negative				
(63) Sysmex	CVP >>>	7	11	Negative				
(109) Axis-Shield	CVP >>>	1	1,5	Negative				
(162) Siemens (Atellica)	CVP >>>	2	3	Negative				
(177) Mindray	CVP >>>	1	1,5	Negative				
(179) Siemens	CVP >>>	10	15	Negative				
(197) Analyticon	CVP >>>	1	1,5	Negative				
(202) Dirui	CVP >>>	19	29	Negative				
(207) 77 Elektronika Kft.	CVP >>>	2	3	Negative				
(999) another manufacturer	CVP >>>	3	4,5	Negative				
<b>Sample B</b>		66				63	61	97
(5) Dialab	>	1	1,5	> 250 leu/μL				
(12) Beckman Coulter; (223) Beckman Coulter iChem Velocity / DxU 810c Iris	±	3	4,5	Negative				
(18) Cormay	>	1	1,5	10 - 50 leu/μL				
	CVP >>>	5	7,6	51 - 250 leu/μL				
(46) Erba Lachema	CVP >>>	1	1,5	51 - 250 leu/μL				
	>	6	9,1	> 250 leu/μL				
(49) BioVendor	CVP >>>	1	1,5	51 - 250 leu/μL				
(60) Roche	CVP >>>	2	3	51 - 250 leu/μL				
(63) Sysmex		2	3	Negative				
	>	3	4,5	10 - 50 leu/μL				
	CVP >>>	2	3	51 - 250 leu/μL				
(109) Axis-Shield	>	1	1,5	> 250 leu/μL				
(162) Siemens (Atellica)	>	2	3	> 250 leu/μL				
(177) Mindray	>	1	1,5	10 - 50 leu/μL				
(179) Siemens	CVP >>>	2	3	51 - 250 leu/μL				
	>	8	12	> 250 leu/μL				
(197) Analyticon	>	1	1,5	10 - 50 leu/μL				
(202) Dirui	>	4	6,1	10 - 50 leu/μL				
	CVP >>>	12	18	51 - 250 leu/μL				
	>	3	4,5	> 250 leu/μL				
(207) 77 Elektronika Kft.	CVP >>>	2	3	51 - 250 leu/μL				
(999) another manufacturer	CVP >>>	1	1,5	51 - 250 leu/μL				
	>	2	3	> 250 leu/μL				
<b>(274) Nitrite</b>		69				69	69	100
<b>Sample A</b>		69				69	69	100
(5) Dialab	CVP >>>	1	1,4	Negative				
(12) Beckman Coulter	CVP >>>	3	4,3	Negative				
(18) Cormay	CVP >>>	6	8,7	Negative				
(46) Erba Lachema	CVP >>>	9	13	Negative				
(49) BioVendor	CVP >>>	1	1,4	Negative				
(60) Roche	CVP >>>	3	4,3	Negative				
(63) Sysmex	CVP >>>	7	10	Negative				
(109) Axis-Shield	CVP >>>	1	1,4	Negative				
(162) Siemens (Atellica)	CVP >>>	2	2,9	Negative				
(177) Mindray	CVP >>>	1	1,4	Negative				
(179) Siemens	CVP >>>	10	14	Negative				
(197) Analyticon	CVP >>>	1	1,4	Negative				
(202) Dirui	CVP >>>	19	28	Negative				
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	Negative				
(999) another manufacturer	CVP >>>	3	4,3	Negative				
<b>Sample B</b>		69				69	69	100
(5) Dialab	CVP >>>	1	1,4	Positive				
(12) Beckman Coulter	CVP >>>	3	4,3	Positive				
(18) Cormay	CVP >>>	6	8,7	Positive				
(46) Erba Lachema	CVP >>>	9	13	Positive				
(49) BioVendor	CVP >>>	1	1,4	Positive				
(60) Roche	CVP >>>	3	4,3	Positive				
(63) Sysmex	CVP >>>	7	10	Positive				
(109) Axis-Shield	CVP >>>	1	1,4	Positive				

## EQA round: DGP2/26 - Urine Strip Tests Analysis

Deadline: 1.5.2026

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N <sub>tot</sub>	Frequency of the results		Success		
			N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
(162) Siemens (Atellica)	CVP >>>	2	2,9	Positive			
(177) Mindray	CVP >>>	1	1,4	Positive			
(179) Siemens	CVP >>>	10	14	Positive			
(197) Analyticon	CVP >>>	1	1,4	Positive			
(202) Dirui	CVP >>>	19	28	Positive			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	Positive			
(999) another manufacturer	CVP >>>	3	4,3	Positive			
<b>(273) pH</b>		70			70	69	99
<b>Sample A</b>		70			70	69	99
(5) Dialab	>	1	1,4	6,0			
(12) Beckman Coulter	CVP >>>	2	2,9	5,0			
	>	1	1,4	5,5			
(18) Cormay	CVP >>>	4	5,7	5,0			
	>	2	2,9	6,0			
(46) Erba Lachema	CVP >>>	6	8,6	5,0			
	>	1	1,4	5,5			
	>	3	4,3	6,0			
(49) BioVendor	>	1	1,4	5,5			
(60) Roche	CVP >>>	3	4,3	5,0			
(63) Sysmex	CVP >>>	7	10	5,0			
(109) Axis-Shield	>	1	1,4	6,0			
(162) Siemens (Atellica)	>	1	1,4	5,5			
	>	1	1,4	6,0			
(177) Mindray	CVP >>>	1	1,4	5,0			
(179) Siemens	CVP >>>	1	1,4	5,0			
	>	7	10	5,5			
	>	1	1,4	6,0			
		1	1,4	7,0			
(197) Analyticon	CVP >>>	1	1,4	5,0			
(202) Dirui	CVP >>>	10	14	5,0			
	>	9	13	5,5			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	5,0			
(999) another manufacturer	CVP >>>	1	1,4	5,0			
	>	2	2,9	6,0			
<b>Sample B</b>		70			70	70	100
(5) Dialab	CVP >>>	1	1,4	7,5			
(12) Beckman Coulter	>	1	1,4	6,5			
	CVP >>>	2	2,9	7,0			
(18) Cormay	>	1	1,4	6,5			
	CVP >>>	2	2,9	7,0			
	CVP >>>	2	2,9	7,5			
	>	1	1,4	8,0			
(46) Erba Lachema	CVP >>>	8	11	7,0			
	CVP >>>	1	1,4	7,5			
	>	1	1,4	8,0			
(49) BioVendor	>	1	1,4	6,5			
(60) Roche	CVP >>>	1	1,4	7,0			
	>	2	2,9	8,0			
(63) Sysmex	CVP >>>	7	10	7,5			
(109) Axis-Shield	CVP >>>	1	1,4	7,5			
(162) Siemens (Atellica)	CVP >>>	2	2,9	7,0			
(177) Mindray	>	1	1,4	8,0			
(179) Siemens	CVP >>>	8	11	7,0			
	CVP >>>	2	2,9	7,5			
(197) Analyticon	>	1	1,4	6,5			
(202) Dirui	CVP >>>	7	10	7,0			
	CVP >>>	12	17	7,5			
(207) 77 Elektronika Kft.	>	2	2,9	6,5			
(999) another manufacturer	CVP >>>	3	4,3	7,5			
<b>(266) Urobilinogen</b>		70			70	68	97
<b>Sample A</b>		70			70	70	100
(5) Dialab	CVP >>>	1	1,4	Negative, < 17 µmol/L			
(12) Beckman Coulter	CVP >>>	3	4,3	Negative, < 17 µmol/L			
(18) Cormay	CVP >>>	6	8,6	Negative, < 17 µmol/L			
(46) Erba Lachema	CVP >>>	10	14	Negative, < 17 µmol/L			
(49) BioVendor	CVP >>>	1	1,4	Negative, < 17 µmol/L			
(60) Roche	CVP >>>	3	4,3	Negative, < 17 µmol/L			
(63) Sysmex	CVP >>>	7	10	Negative, < 17 µmol/L			
(109) Axis-Shield	CVP >>>	1	1,4	Negative, < 17 µmol/L			

## EQA round: DGP2/26 - Urine Strip Tests Analysis

Deadline: 1.5.2026

Setup: groups - R (strip manufacturer)

Test Sample Group	Frequency of the results				Success		
	AV	N <sub>tot</sub>	N <sub>rel</sub> [%]	Result	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub> [%]
(162) Siemens (Atellica)	CVP >>>	2	2,9	Negative, < 17 µmol/L			
(177) Mindray	CVP >>>	1	1,4	Negative, < 17 µmol/L			
(179) Siemens	CVP >>>	10	14	Negative, < 17 µmol/L			
(197) Analyticon	CVP >>>	1	1,4	Negative, < 17 µmol/L			
(202) Dirui	CVP >>>	19	27	Negative, < 17 µmol/L			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	Negative, < 17 µmol/L			
(999) another manufacturer	CVP >>>	3	4,3	Negative, < 17 µmol/L			
<b>Sample B</b>		70			70	68	97
(5) Dialab	>	1	1,4	1+, 2+			
(12) Beckman Coulter	>	3	4,3	1+, 2+			
(18) Cormay	CVP >>>	6	8,6	3+, 4+			
(46) Erba Lachema	>	4	5,7	1+, 2+			
	CVP >>>	6	8,6	3+, 4+			
(49) BioVendor	CVP >>>	1	1,4	3+, 4+			
(60) Roche	CVP >>>	3	4,3	3+, 4+			
(63) Sysmex	>	3	4,3	1+, 2+			
	CVP >>>	4	5,7	3+, 4+			
(109) Axis-Shield	CVP >>>	1	1,4	3+, 4+			
(162) Siemens (Atellica)	>	2	2,9	1+, 2+			
(177) Mindray	CVP >>>	1	1,4	3+, 4+			
(179) Siemens		2	2,9	Negative, < 17 µmol/L			
	>	4	5,7	1+, 2+			
	CVP >>>	4	5,7	3+, 4+			
(197) Analyticon	CVP >>>	1	1,4	3+, 4+			
(202) Dirui	>	5	7,1	1+, 2+			
	CVP >>>	14	20	3+, 4+			
(207) 77 Elektronika Kft.	CVP >>>	2	2,9	3+, 4+			
(999) another manufacturer	CVP >>>	3	4,3	3+, 4+			